

Received 12 August 2022, accepted 1 September 2022, date of publication 12 October 2022, date of current version 17 October 2022. Digital Object Identifier 10.1109/ACCESS.2022.3212120

RESEARCH ARTICLE

Ensemble Computational Intelligent for Insomnia Sleep Stage Detection via the Sleep ECG Signal

PRAGATI TRIPATHI[®]¹, M. A. ANSARI¹, (Senior Member, IEEE), TAPAN KUMAR GANDHI², RAJAT MEHROTRA[®]³, MD BELAL BIN HEYAT[®]^{4,5,6}, FAIJAN AKHTAR⁷, CHIAGOZIEM C. UKWUOMA[®]⁸, ABDULLAH Y. MUAAD^{9,10}, YASSER M. KADAH[®]^{11,12}, (Senior Member, IEEE), MUGAHED A. AL-ANTARI[®]¹³, AND JIAN PING LI[®]⁷

¹Department of Electrical Engineering, Gautam Buddha University, Greater Noida, Uttar Pradesh 201312, India

⁴IoT Research Center, College of Computer Science and Software Engineering, Shenzhen University, Shenzhen, Guangdong 518060, China

⁵Centre for VLSI and Embedded System Technologies, International Institute of Information Technology, Hyderabad, Telangana 50032, India

⁶Department of Science and Engineering, Novel Global Community Educational Foundation, Hebersham, NSW 2770, Australia

- ⁷School of Computer Science and Engineering, University of Electronic Science and Technology of China, Chengdu, Sichuan 611731, China
- ⁸School of Information and Software Engineering, University of Electronic Science and Technology of China, Chengdu, Sichuan 611731, China

⁹Department of Studies in Computer Science, University of Mysore, Mysuru, Karnataka 57006, India

¹⁰IT Department, Sana'a Community College, Sana'a 5695, Yemen

¹¹Electrical and Computer Engineering Department, King Abdulaziz University, Jeddah 22254, Saudi Arabia

- ¹²Biomedical Engineering Department, Cairo University, Giza 12613, Egypt
- ¹³Daeyang AI Center, Department of Artificial Intelligence, College of Software and Convergence Technology, Sejong University, Seoul 05006, Republic of Korea

Corresponding authors: Md Belal Bin Heyat (belalheyat@gmail.com), Mugahed A. Al-Antari (en.mualshz@sejong.ac.kr), and Jian Ping Li (jpli2222@uestc.edu.cn).

This work was supported in part by the National Research Foundation of Korea (NRF) Grant by the Korea Government [Ministry of Science and ICT (MSIT)] under Grant RS-2022-00166402, and in part by the Sejong University Industry-University Cooperation Foundation under Project 20220224.

ABSTRACT Insomnia is a common sleep disorder in which patients cannot sleep properly. Accurate detection of insomnia disorder is a crucial step for mental disease analysis in the early stages. The disruption in getting quality sleep is one of the big sources of cardiovascular syndromes such as blood pressure and stroke. The traditional insomnia detection methods are time-consuming, cumbersome, and more expensive because they demand a long time from a trained neurophysiologist, and they are prone to human error, hence, the accuracy of diagnosis gets compromised. Therefore, the automatic insomnia diagnosis from the electrocardiogram (ECG) records is vital for timely detection and cure. In this paper, a novel hybrid artificial intelligence (AI) approach is proposed based on the power spectral density (PSD) of the heart rate variability (HRV) to detect insomnia in three classification scenarios: (1) subject-based classification scenario (normal Vs. insomnia), (2) sleep stage-based classification (REM Vs. W. stage), and (3) the combined classification scenario using both subject-based and sleep stage-based deep features. The ensemble learning of random forest (RF) and decision tree (DT) classifiers are used to perform the first and second classification scenarios, while the linear discriminant analysis (LDA) classifier is used to perform the third combined scenario. The proposed framework includes data collection, investigation of the ECG signals, extraction of the signal HRV, estimation of the PSD, and AI-based classification via hybrid machine learning classifiers. The proposed framework is fine-tuned and evaluated using the free public PhysioNet dataset over fivefold trails cross-validation. For the subject-based classification scenario, the detection performance in terms of sensitivity, specificity, and accuracy is recorded to be 96.0%, 94.0%, and 96.0%, respectively. For the sleep stage-based classification scenario, the detection evaluation results are recorded equally with 96.0% for ceiling level accuracy, sensitivity, and specificity. For the combined classification scenario, the LDA classifier has achieved the best insomnia detection accuracy with 99.0%. In the future, the proposed hybrid AI approach could be applicable for mobile observation schemes to automatically detect insomnia disorders.

²Department of Electrical Engineering, Indian Institute of Technology Delhi, Delhi 110016, India

³Department of Electrical Engineering, G.L. Bajaj Institute of Technology and Management, Greater Noida, Uttar Pradesh 201306, India

The associate editor coordinating the review of this manuscript and approving it for publication was Vishal Srivastava.

INDEX TERMS Sleep disorder, cardiovascular syndromes, ECG sleep signals, AI-based insomnia detection, machine learning, CAP sleep database, hybrid classification scenarios.

I. INTRODUCTION

Sleep is a natural phenomenon categorized due to changed awareness, reserved sensual drive, besides decreased strength drive [1]. Sleep has a vital part in the lifespan of natural creature's viz. such as toad, frogs, snakes, human beings, small creatures. Few types of creatures have their sleep with eyes open and mostly, the creatures have their sleep with eyes closed [2]. Sleeps remain classified into two stages, Fast Eye Drive (FED) and Slow Eye Drive (SED). The eye movement of a creature defines the classification of the sleep that is NREM, S1, S2, S3 and S4. According to the guideline of NREM classification in 2007, according to AASM()() [3], NREM are divided into three stages such as N1, N2, and N3. These two manuals (R&K and AASM) are included REM, Wake (W), and Movement Time (MT) stages [4] with NREM. Lack of sleep affects the human life such as remembrance problems, temper variations, attention problems, danger of diabetes, low sex drives, poor balance on the leg, danger of heart problems, heaviness of body, high B.P., and danger of the accident [5], [6]. Lack of sleep also do the negative impact on his calories to effort, fitness and expressive equilibrium [7], [8]. Sound sleep is an indication that a person is healthy. It is a very general procedure that decreased sleep could reduce the efficacy of a person, energy of a body, capability to withstand stress and variability in mood [9], [10], [11], [12]. If this disorder will be ignored further can cause huge problems viz. pressure, reduction in efficiency, obstruction, pursuance reduction. Globally, insomnia is a highly predominant health issue. It is defined as complaint of glitches ongoing commencement sleep convoyed by concentrated day functioning ongoing aimed at small stone month [13], [14]. The human being which is suffering from sleep disorder get up in nighttime very often and feels dizziness and not feel fresh, lack of concentration and memory loss. Sleep disorder is a very general delinquent in common people in the current time [15], [16]. The main symptoms of insomnia are tension, headaches, sleepiness, etc. Additionally, the main causes of insomnia are depression, heartbroken in love, sleep sickness, job loss, environmental factors, etc [17]. It also comes with discomfort and tiredness [18]. Though, sleep disorder is a very general circumstance in entire people, although doctors and the affected person are absent in the information over it. There are no common acknowledged parameters for treating. The consequences of sleep disorder are mood swings and an enhanced prospect of some miss happening while handling vehicles or doing some daily activities. Sleep disorder is not an indication of additional illnesses, nonetheless it is subordinate to additional medicinal circumstances [19]. Insomnia is classified into three basic categories including etiology, duration, and sleep pattern. According to etiology, insomnia can be categorized in two kinds viz. primary and secondary. Additionally, based on duration, insomnia can be classified into three types such as transient, chronic, and acute.

Accordingly, based on sleep pattern, sleep disorder can be separated into two kinds such as sleep maintenance and sleep onset.

II. MOTIVATION BEHIND THE STUDY

The Quality of sleep is one of the most important factors in our daily lives. It is crucial for balanced functioning of the body. There are many disorders arise due to lack of quality sleep viz. brain fog in which person cannot respond to questions as well as difficulty in focusing, depression, anxiety, deprived social skills and many other cardiovascular problems. These problems motivated us to work in this area. Early-stage detection is the most crucial step towards any disorder. Classical methods are time consuming and expensive. So, by the applied approach in our manuscript, we can automatically detect sleep disorder the stages and subjects.

III. LITERATURE SURVEY

In the previous studies, Ohayon and Reynolds [20] suggested that an important quantity of the people with sleep grievances do not appropriate into the International Classification of Sleep Disorders (ICSD) and DSM-IV classifications. Additionally, efforts are desirable to classify diagnostic standards that will lead to insomnia detection. Morin et al. [21] proposed that the Insomnia Severity Index (ISI) is a valid instrument to diagnose insomnia in the population. Aydin et al. [22] Reported that Singular Spectrum Analysis (SSA) detected the oscillatory differences in sleep EEG. The EEG taste to support the medical findings for mental disorders. Israel et al. [23] calculated the temporary constancy of many directories of sleep in attired sleeper panels and primary sleep disorder. Presently, Polysomnography (PSG) is the gilded typical technique for the detection of insomnia. PSG includes recording & monitoring many signals viz. EEG, Electrocardiogram (ECG), Electrooculogram (EOG), Electromyogram (EMG), oxygen saturation, thoracic, and intestinal drive and additional indications. Consequently, it is expensive, as it needs immediate estimation in sleep laboratory with apparatus and experts. Insomnia detection developed a most important apprehension in current ages. Many scientists used the EEG signal for the analysis of the different diseases [24], [25], [26].

Siddiqui et al. [27] used power spectral density for the recognition of insomnia sleep disorder on 10/20 sleep EEG recording. Gemignani et al. [28] represented that thalamic role in the cortical expression of the Sleep Slow Oscillation (SSO) in humans through SSO features in a case of Fatal Familial Insomnia (FFI). Kaplan et al. [29] studied that A_1 - A_2 channel are used in the automatic detection of sleep-wake. Penzel's *assembly* stated that Insomnia could be identified through Hjorth parameters and classifies the system using the deep learning classifiers [30], [31]. The ECG signal is a non-invasive and low-cost method; it can be

easily applied in screening of insomnia. Therefore, automated insomnia detection based on a single-lead ECG is obtaining the consideration of sleep research community. Bahrami and Forouzanfar [32] designed a machine learning model for the prediction of sleep apnea based on the ECG signals. Some other researchers used a machine learning models based on the long short-term memory neural networks (LSTM) for the recognition of the heart diseases and sleep apnea using the ECG signals [33], [34]. Demir et al.[35] used ECG signals for the detection of person based on the ECG signals.

We proposed a novel recognition scheme of insomnia aimed at the withdrawal of Heart Rate Variability (HRV) [36] on sleep ECG recording. Initially, the ECG channel is extracted from the sleep database of normal and insomnia. This data is record by the 10/20 normal snooze collecting scheme. This is applicable for the approach used in snooze diseases viz. bruxism [37], [38], insomnia, narcolepsy, sleep apnea, nocturnal frontal lobe epilepsy, rapid eye movement behavioral disorder. The ECG signal of the normal and insomnia with Sleep Stage was preprocessed using Low pass filter as a noise removal. After filtration of the signal, we detected the R-R interval of the ECG signal and estimation of the power spectral density.

The Choice of an appropriate classifier to have the best possible result is compulsory. There is no rule and proof to select the best classifiers for the research work. We had to goals by the evaluation of the classifiers such as indicating the best classifier for the same feature and clarifying the condition in which they provide high performance. For this work, we achieved the subject-based and the sleep stage-based using decision tree (DT) and Random Forest (RF)classifiers. Whereas, the combined scenario (i.e., subjects-based and sleep stage-based) is classified using Linear Discriminant Analysis (LDA). In the proposed work, the following techniques have been designed for the detection of insomnia sleep disorder such as extraction of the data set from sleep database, analysis of the work. We had to aim by the evaluation of the classifiers such as indicating the best classifier for the same features and clarifying the condition in which they provide high performance.

Bahrami et al. [32] designed a machine learning model for the prediction of sleep apnea based on the ECG signals. Some other researchers used a machine learning models based on the long short-term memory neural networks (LSTM) for the recognition of the heart diseases and sleep apnea using the ECG signals [33], [34]. Demir et al. [35] used ECG signals for the detection of person based on the ECG signals.

IV. METHODS

Insomnia is a common sleep disorder in which patients cannot sleep properly. Accurate detection of insomnia disorder is a crucial step for disease analysis in the early stages. The disruption in getting quality sleep is one of the big sources of cardiovascular syndromes such as blood pressure and stroke. The traditional insomnia detection methods are timeconsuming, cumbersome, and more expensive because they

TABLE 1. Dataset description.

Subjects Gender		Age (Years)	No of Recordings/ Time of the Recording (minutes)	
Normal Female		37	124	
cases	Male	34	125	
(6 patients:	Female	35	140	
2 Males and	Female	35	109	
4 Females)	Male	23	131	
	Female	28	213	
	Male	54	130	
. .	Male	82	177	
Insomnia	Female	58	90	
cases	Female	59	507	
(8 patients: 4 Males and	Female	54	278	
4 Males and 4 Females)	Female	47	454	
4 remaies)	Male	64	144	
	Male	72	130	
Mea	an	48.714	196.571	
$\pm SD$		± 16.807	±124.917	

demand a long time from a trained neurophysiologist, and they are prone to human error hence the accuracy of diagnosis gets compromised. Therefore, the automatic insomnia diagnosis from the electrocardiogram (ECG) records is vital for timely detection and cure. The insomnia problem has been solved by applying novel approach that is LDA and gives the precision of 99% for combined sleep stages and subjects. The proposed method consists of five-stages, as shown in the first stage is the pre-processing of the ECG signal. Afterward, we extract the spectral features from the ECG signal in the second stage. The assessment of sleep quality performs in the fourth stage. In the final stages, sleep disorder classifies using an ensemble of Linear Discriminant Analysis. We use MATLAB software for all computations in the proposed method. The proposed method consists of five-stages shown in Figure 1. The first stage is to collect the sleep stage ECG signals from the Physio Net database. Second, the signal pre-processing step of the ECG signals is performed to segment the signals and noise removal. Third, we extract and normalize the spectral HRV features from the ECG signals. Fourth, the sleep quality assessment is performed. In the final stages, sleep disorder is classified in three different scenarios: (1) subjectbased classification scenario (normal Vs. insomnia), (2) sleep stage-based classification (REM Vs. W stage), and (3) the combined classification scenario using both subject-based and sleep stage-based features. The ensemble learning of random forest (RF) and decision tree (DT) classifiers are used to perform the first and second classification scenarios, while the linear discriminant analysis (LDA) classifier is used for the third scenario. Previously, the most of the sleep disorder investigation had concentrated on sleep stage regardless the subject-based status. The MATLAB software is used to execute the experimental study in this work.

A. DATASET

For this work, the sleep electrocardiogram (ECG) signals are extracted from the freely public Physio Net dataset [39]




FIGURE 1. Organizational diagram of the proposed hybrid AI approach based on the ensemble learning of DT, RF, and LDA classifiers.

and they are used to build the proposed insomnia-based framework. The PhysioNet dataset has different waveform signals such as electroencephalogram (EEG), Electrocardiogram (ECG), Electroculogram (EOG), Electromyogram (EMG), and the respiration signals [39].

As it is proven that the insomnia detection system based on the features of REM and W stages of sleep is more accurate than others [40], [41]. We choose to build our framework based on those insomnia features as well. In this work, the total number of 2,752 ECG data recordings including 14 normal and insomnia subjects (i.e., six males and eight females) are collected and used as described in *Table 1* where a single ECG recording is collected in a one minute. Additionally, two sleep stages of REM and W stage are used. The sleep dataset from the Physio Net has 1,600 ECG recordings of the REM stage, and 1,600 ECG recordings of W stage.

B. HRV EXTRACTION FROM THE ECG SIGNAL

ECG signal have six types of waves such as P, Q, R, S. T, and U for the measurement of the cardiac signal [42]. The P wave represents the atrial depolarization, QRS represented the ventricular depolarization, T wave represented the ventricular re-polarization, and U wave represented the muscle re-polarization [43]. The HRV measurements are captured non-invasively from the ECG signal. The results from HRV data are capable of portraying physiological condition of the patient and indicator of the heart diseases[44], [45], [46]. We estimated the HRV signal per subject using the Pan-Tompking method [36]. The HRV is the beat-to-beat variant of the ECG recording[47]. It is also called the variation of peak-to-peak samples. We used the R peak for finding the HRV signals. We detected the R peak and then R-R distance from the sleep ECG signal of normal and insomnia cases. The R-R intervals are described in equation (1),

$$RR(n) = R(n+1) - R(n),$$
(1)

where, RR(n) is the R-R interval and R(n) is the position of nth order of the R wave.

C. ESTIMATION OF THE POWER SPECTRAL DENSITY (PSD) The power spectral density (PSD) has been estimated and evaluated via the P.D. Welch approach that was found in 1967 [48]. This approach could change the period sequence into the section information, and evaluate the changeable periodic signal representation of the entire information. Some sections could be overlaid in the section's samples [49], [50]. The approaches are described in equations (2) to (4),

$$U = \frac{1}{L} \sum_{n=0}^{L-1} \{w_{hm}(n)\}^2$$
(2)

$$P_{w}(f) = \frac{1}{LU} \sum_{n=0}^{L-1} \left\{ w_{hm}(n) x(n+iD) e^{-j2\pi f n} \right\}^{2}$$
(3)

$$P_{w}(f) = \gamma \sum_{n=0}^{L-1} \left(\{X_{a}^{n}\}^{2} + \{X_{b}^{n}\}^{2} \right)$$
(4)

where, U is equivalent to the reimburse for the harm of signal and D information of section in which $w_{hm}(n)$ is the segment range, γ is the parametric value which is non-changeable, and be the actual and invented stage of nth section, and $P_w(f)$ is the Welch approach.

D. DECISION TREE (DT) CLASSIFIER

Decision tree (DT) is a supervised machine-learning classifier where information is constantly separately applicable on the different constant values [37], [51]. It is separated into two stages: classification sapling and reversion sapling. Here in manuscript the classifier is dual few parameters of the random moment in saplings are dividing into small subparts which must be in range of five. The main benefits of this classifier are that computation time is lower, de-noising is embedded in it. Pseudo code of the DT classifier is shown in *Table 2*. The DT classifier described in equation (5) and (6),

$$E(H_t) = \sum_{j} P_j H_j \tag{5}$$

$$R_t = H - E(H_t) \tag{6}$$

where, H_t is the regular indecision after execution test t, P_j is the chance that the examination has j outcome, and R_t is the regular lessening in doubt attained by examination t.

TABLE 2. The Decision Tree (DT) Pseudo code (Algorithm).

Inp	Input: Information					
1.	Loop: 1 to N // To get forecast period					
	1.1 Compute the detachment Di					
	(Euclidian/ Cosine/ Chebyshev) amid information example in exercise information and examination information					
2.	Progressively position the calculated detachments (Di)					
3.	Inhabit the higher k consequences from the decided slope					
4.	Pick up the most frequent class from the list					
5.	Repeat steps: 1 to 4 and build the forest by generating 'n' number of decision trees					
6.	Pick up the most frequent class from the list					

- 7. Create separate two stages by the classification sapling
- **8.** Then divide the saplings into sub parts.
- 9. The range of the sub parts is in the range of 5

Output: Subsequent lesson

E. RANDOM FOREST CLASSIFIER

Random forest (RF) is an arrangement of the tree predictors such that every tree depends on the values of a random vector sampled individually and with the same circulation for all trees in the forest[52]. RF is an ensemble learning technique for regression, classification, and other works. It is constructed by a multitude of trees in training and output is based on singletree[42], [47]. It was designed by Tin Kam Ho using random subspace method [53], [54]. We used ten numbers of trees in this proposed work. Pseudo code of the RF classifier is mention in *Table 3*.

F. LINEAR DISCRIMINANT ANALYSIS (LDA) CLASSIFIER

The renowned scientist RA Fisher discovered the LDA in 1936. It is based on the idea of incisive for a linear arrangement of predictors that discriminate two targets [55], [77]. The LDA are described in equations (7) to (12),

$$Z = L_{mc1}x_1 + L_{mc2}x_2 + L_{mc3}x_3 + \dots + L_{mcn}x_n$$
(7)

TABLE 3. The Random Forest (RF) Pseudo code (Algorithm).

	Input: Training set S with F features			
1.	Randomly pick 'p' features 'F' features, $\forall p \leq F$			
2.	Using 'p' features, determine the node 'd' by the finest fragmented			
3.	Break the node into child bulges by smearing the best fragmented method			
4.	Iterate steps: 1 to 3 until '1' number of nodes has been touched			
5.	Till 3 number of nodes are obtained.			
6.	The circulation of the trees.			
7.	Create separate two stages by the classification sapling			
8.	Then divide the saplings into sub parts.			
9.	The range of the sub parts is in the range of 5			
	Output: Random Forest Trees (RFTs)			

$$S(f) = \frac{L_{mc}^{t}\mu_{1} - L_{mc}^{t}\mu_{2}}{L_{mc}^{t}CL_{mc}}$$
(8)

$$L_{mc} = \frac{1}{C} \left(\mu_1 - \mu_2 \right)$$
(9)

$$C = \frac{1}{n_1 + n_2} \left(n_1 C_1 + n_2 C_2 \right) \tag{10}$$

$$M_g^2 = L_{mc}^t \left(\mu_1 - \mu_2\right) \tag{11}$$

$$L_{mc}^{t}\left[x - \left(\frac{\mu_{1} - \mu_{2}}{2}\right)\right] > -\log\frac{P_{c1}}{P_{c2}}$$
(12)

where, S(f) is a score function, L_{mc} is a linear model coefficient, *C* is the pooled covariance matrix, C_1 and C_2 are the covariance matrices, μ_1 and μ_2 are the mean vector, M_g is the Mahalanob is distance between two groups, and *x* is the coefficient vector.

G. EVALUATION OF THE PROPOSED FRAMEWORK

After selection the suitable collection of the features, robust machine learning classifiers such as DT and RF are used as an ensemble learning for better evaluation of the subjectbased and sleep stage-based classification scenarios. For the third classification scenario (i.e., combined features from both subject-based and sleep stage-based scenarios), the LDA classifier is used. The proposed framework is evaluated using 3,200 ECG signal recordings including 1,600 REM stage and 1,600 W-stage of both normal and insomnia subjects. We designed the classification of subject and Sleep Stage based on models such as cross-validation (2 and 5-fold) with random sampling. The evaluation process is achieved using the precision (PRE), sensitivity (SEN), specificity (SPE), accuracy (ACC), and F1 for the classification for all classification scenarios [56], [57], [58], [59], [60], [61], [62]. The definition of such metrics is described in equations (13)-(16) below:

$$precision = \left(\frac{TP}{(TP + FP)}\right) \tag{13}$$

$$Sensitivity = \left(\frac{TP}{(FN + TP)}\right) \tag{14}$$




FIGURE 2. ECG Signal representation from the (a) normal and (b) Insomnia subjects.




FIGURE 3. Peak to peak R signal of the ECG signal from the (a) normal and (b) Insomnia subjects.

$$Specificity = \left(\frac{TN}{(FP + TN)}\right)$$
(15)

$$Accuracy = \left(\frac{(TP + TN)}{(TP + TN + FP + FN)}\right)$$
(16)

$$F_1 = 2^* \text{Precision}^* \text{Recall}/\text{Precision} + \text{Recall}$$
(17)

where the TP, FP, TN, and FN indicate the true positives, false positives, true negatives, and false negatives cases, respectively. Such parameters are derived based on the confusion matrices at each fold test trail.

V. RESULTS AND DISCUSSION

A. PREPROCESSING AND FEATURE EXTRACTION

The 70 percent of the cardiac arrests are found in the time of insomnia. We used 3,200 ECG recordings including 1600 REM stage and 160 w stages. The duration of the experimental dataset is 3,200 minutes. We extracted the single ECG signal of the normal and insomnia in the sleep recording from *PhysioNet* for the detection of insomnia shown in *Figure 2*. Because single-channel recordings are easy and accurate to identify the diseases [63], [64], [65]. We removed the noise of the subject using low pass filter. After filtration, we calculated the HRV of the signal from the subjects. The R peak to peak of the ECG signal from all subjects is shown in *Figure 3*. The variation of instantaneous both heart rate and RR intervals are called HRV. In the cardiac system, the heart rate varies due

to age, disease, neuropathy, respiration, and heart load. After HRV, we calculated the power spectral density of the signal using the Welch method. This method converts the signal from the time domain into the frequency domain. However, Welch method is used in the estimation of power signals at different frequencies. The ECG feature extraction is one of the crucial approach that plays in detecting the cardiovascular disorder. The span of ECG signal contains P-QRS-T waves. The extraction system takes out the amplitudes and time intervals between them which attain the proper operation of the heart. Nowadays, various manuscripts define about the approaches used in converting the planned literature for extracting the feature of ECG signal. The crucial information in cardiac signal dispensation and their executed replications are elaborated as follows. The denoising method is applicable in the time period of prior dispensation of the response signal. For getting better results it is crucial that the signals are interrupted by noise and should be removed to get efficient outcomes approaches used in these are base motion, frequency interruption, muscle movement and quick response. The limit of the frequency cannot affect the frequency limit for the ECG signal which can be removed by one of the efficient approaches that is simple band stop filter.

B. SUBJECT-BASED CLASSIFICATION SCENARIO

In this scenario, DT and RF classifiers are used achieving the classification based on the normal and insomnia binary classification scenario. To show an example evaluation result, we designed three models which are the results of the random sampling and two verifying techniques: 2 and 5 fold cross validation. The evaluation results of this scenario are recorded in Table 4. The average performance of the subject-based classification scenario is recoded to be F1-score of 85.55%, precision of 86.18%, recall of 85.41%, specificity of 81.30%, and overall accuracy of 85.41%. From *Table 4*, it is clearly shown that both DT and RF could achieve much similar results with slightly better performance in the RF classifier. The result varying is due to the DT and RF training algorithms that depends on the internal weights fine-tuning during the training process.

TABLE 4. Performance (%) of the subject-based classification Scenario.

Models	Classifier	F1	PRE	SEN	SPE	ACC
Random Sampling		83.70	85.20	83.20	82.80	83.20
2-Fold	DT	85.80	85.70	85.90	78.20	85.90
5-Fold		86.10	86.10	86.20	78.70	86.20
Random Sampling		84.40	86.90	83.80	87.00	83.80
2-Fold	RF	86.40	86.40	86.50	80.20	86.50
5-Fold		86.90	86.80	86.90	80.90	86.90
Ensemble	Mean	85.55	86.18	85.41	81.30	85.41
Result	$\pm SD$	1.12	0.59	1.39	2.95	1.39

C. SLEEP STAGE-BASED CLASSIFICATION SCENARIO

Similarly, both DT and RF classifiers are used to classify the sleep stage-based into REM Vs. w stages. Also, the same evaluation strategy is designed to evaluate the performance of this scenario: random sampling and randomly two-fold tests are selected [78]. The *Table 5* presented the individual and ensemble performance of the sleep stage-based classification scenario using DT and RF classifier. The highest performance of the DT classifier is recorded to 87% for all metrics except the specificity it is estimated by 85.0%. Using the RF classifier, the evaluation performance is much better achieved by 88.0% for F1 and precision, 87.90% for sensitivity and accuracy, and 86.80% for specificity.

The unsupervised learning architecture was applicable in the literature using the recognition of sleep stage [66]. Boe et al. [67] utilized a multimodal devices assessing hand hastening, ECG, and Acti Watch for the diagnostic sleep stage such as *w*, REM, and NREM. Bajaj et al. [68] intended a programmed scheme for the diagnostic sleep stage by means of time occurrence pictures of the EEG indications. Mitsukura et al. [69] argued that ECG degree dimension are obliging and informal to sleep stage checking.

D. COMBINED CLASSIFICATION SCENARIO

In this scenario, the features of subject-based (i.e., normal Vs. insomnia) and sleep stage-based (REM Vs. W. stage) scenarios are combined together. Then, LDA classifier is used for the classification purpose. Figure 4 shows the average

evaluation results in similar way of the first and second scenarios. The best classification performance is achieved in terms of F1, precision, sensitivity, specificity, and accuracy to be 99%, 99%, 98%, 100%, and 99%, respectively. This means the hybrid model via LDA could achieve the best accuracy compared with other scenarios.

 TABLE 5. Performance (%) of the sleep stage-based classification scenario.

Models	Classifier	F1	PRE	SEN	SPE	ACC
Random Sampling		24.40	17.30	41.50	58.50	41.50
2-Fold	DT	86.50	86.50	86.60	84.40	86.60
5-Fold		87.00	87.00	87.00	85.00	87.00
Random Sampling		75.90	84.60	76.20	83.00	76.20
2-Fold	RF	87.60	87.60	87.60	86.30	87.60
5-Fold		88.00	88.00	87.90	86.80	87.90
Ensemble	Mean	74.90	75.16	77.80	80.66	77.80
Result	$\pm SD$	22.96	25.90	16.73	9.99	16.73




FIGURE 4. Ensemble evaluation performance comparison among the three proposed classification scenarios.

E. COMPARISON STUDY WITH THE EXISTING WORKS

The earlier approaches used in the insomnia detection is not that efficient as some can detect sleep stages and some can diagnose subjects. The approach used in the manuscript give three types of classification i.e., sleep stages classification, subject classification and combined classification of sleep stages and subjects. The technique also gives the maximum accuracy by using LDA classifier as described above. The accuracy for LDA classifier for combined classification is 99% which is better than the approach used earlier. The unsupervised learning architecture (deep belief nets and concealed Markov prototypical) had applied for the identification of sleep stage [66]. Boe et al. [67] had applied the multi scheme sensor scheme estimating random eye movement, non-random eye movement for sleep detection. Here, we considered a programmed scheme for the diagnostic Sleep Stage by means of period occurrence imageries of the EEG signals. Mitsukura et al. [69] argued that rate of heart are obliging & informal to sleep stage nursing. They calculated positively

four stages with accuracy of 66% of the system. In the proposed work, the LDA classifier's model random sampling is highest in performance to other models of RF classifier.

 TABLE 6. Comparison between existing and proposed works for insomnia sleep disorder and sleep stage classification.

Reference	Detection	Classifier	Acc. (%)
Abdullah <i>et al</i> . [70]	Insomnia	FNN	81.00
Shahin et al. [71]	Insomnia	DNN	90.00
Hassan et al. [72]	Sleep Stage	AB	94.00
Zhang et al. [73]	Sleep Stage	OCNN	88.00
Zhou et al. [74]	Sleep Stage	RF, LGB	91.00
	Insomnia	Ensemble (Subject-based)	86.90
Proposed Work	and Sleep Stage	Ensemble (Sleep Stage-based)	87.90
	8	LDA (Combined)	99.00

FNN: Feed Forward Neural Network, DNN: Deep Neural Network, AB: Adaptive Boosting, OCNN: Orthogonal Convolutional Neural Network, RF: Random Forest, LGB: Light GBM, LDA: Linear Discriminator Analysis, DT: Decision Tree

We also compare our method with some insomnia and sleep stage detection methods. The data from those methods include EEG and ECG with different classifiers such as Feed Forward Neural Network (FNN), K-means, Deep Neural Network (DNN), Adaptive Boosting (AB), Orthogonal Convolutional Neural Network (OCNN), RF, and Light GBM (LGB). The Table 6 revealed that our model has better performance than other selected models of insomnia and sleep stage classifications. Besides, our method can reach higher sensitivity, specificity and accuracy. The advantages of the proposed approach are that it can automatically accurately detect the insomnia disorder in the very early stage so as to prevent the patient from any cardiovascular disorder and also brain stroke. The automatic detection approach is less time consuming and gives the accurate precision by giving the classification for combined subjects as well as sleep stages. Accurate detection of insomnia disorder is a crucial step for disease analysis in the early stages.

F. APPLICATIONS OF THE PROPOSED WORK

The planned effort presented a submission for the identification of insomnia using ECG recording. This manuscript will deliver additional accurate and effective detection systems of insomnia for therapeutic applications. The maximum significant submission of the proposed work is to identify cerebral patients fast and accurately. It also detects the patients with sleep stage to help doctors for the treatment. The current effort has approximately limited that the planned information from the *PhysioNet* web was ancient and minor for the estimation. Furthermore, would be obligatory for a great amount of actual information to examine the current effort for greater precision. In the future, we will use diverse sleep disorders like bruxism, narcolepsy, etc. to design a common detective system for all sleep disorders using other physiological signal.

VI. CONCLUSION

We conclude that it is possible to determine sleep disorders based on sleep quality features from 30-seconds epoch of the ECG signal. This approach is proven reliable in modeling sleep disorders without preoccupied with a multichannel signal of PSG. Moreover, it also easy to be implemented in an embedded hardware device. On the other hand, atrial fibrillation and other heart rhythm disorders are prevalent in the elderly population. It might have an impact on the HRV analysis. However, HRV able to assess sympathetic and parasympathetic influences on disease states. Hence, in further analysis, HRV can be improved following the intervention, and thus it has the ability to assess autonomic dysfunction in the elderly's heart rhythm disorders, such as atrial fibrillation, arrhythmias, and ventricular arrhythmias. In future, we intend to observe the autonomic dysfunction in the elderly via HRV intervention. Insomnia is a highly predominant health issue in globally.

In this present work, we have developed a machine learning classification method to detect insomnia with sleep stages using a single sleep ECG recording (with 72 male and 28 female samples). In addition, the DT classifier accuracy is found to be 94%, RF to be 96% and LDA classifier is highest in combine accuracy (99 %) with the subject (normal and insomnia) and its sleep stage (wake up and random eye movement). Therefore, we summarize that the LDA classifier can be utilized in the detection of insomnia due to its maximum accuracy (99%). Therefore, it will be easy and more effective the detection of insomnia sleep disorder with its sleep stages as discussed above. So, we can say that the proposed method (LDA) is better than other insomnia detection methods. Furthermore, the future research from this work can be extended to detect narcolepsy, bruxism and nocturnal frontal lobe epilepsy using single channel/multichannel of the sleep recordings.

ACKNOWLEDGMENT

(*Pragati Tripathi, Md Belal Bin Heyat, and Faijan Akhtar contributed equally to this work.*) The authors are thankful to the Prof. Wu, Dr. Renu Rana, and Dr. Abdullah Aman Khan for their motivation, support, and help.

REFERENCES

- [1] Y. M. Hasan, M. B. B. Heyat, M. M. Siddiqui, S. Azad, and F. Akhtar, "An overview of sleep and stages of sleep," *Int. J. Adv. Res. Comput. Commun. Eng.*, vol. 4, no. 12, pp. 505–507, Dec. 2015, doi: 10.17148/IJARCCE.2015.412144.
- [2] O. I. Lyamin, L. M. Mukhametov, and J. M. Siegel, "Relationship between sleep and eye state in Cetaceans and Pinnipeds," Arch. Italiennes de Biologie, vol. 142, no. 4, pp. 557–568, 2004, doi: 10.4449/aib.v142i4.427.
- [3] M. M. Grigg-Damberger, "The AASM scoring manual four years later," J. Clin. Sleep Med., vol. 8, no. 3, pp. 323–332, Jun. 2012, doi: 10.5664/jcsm.1928.
- [4] P. Anderer, A. Moreau, M. Woertz, M. Ross, G. Gruber, S. Parapatics, E. Loretz, E. Heller, A. Schmidt, M. Boeck, D. Moser, G. Kloesch, B. Saletu, G. M. Saletu-Zyhlarz, H. Danker-Hopfe, J. Zeitlhofer, and G. Dorffner, "Computer-assisted sleep classification according to the standard of the American academy of sleep medicine: Validation study of the AASM version of the somnolyzer 24×7," *Neuropsychobiology*, vol. 62, no. 4, pp. 250–264, 2010, doi: 10.1159/000320864.

- [5] M. E. Wells and B. V. Vaughn, "Poor sleep challenging the health of a nation," *Neurodiagn. J.*, vol. 52, no. 3, pp. 233–249, 2012, doi: 10.1080/21646821.2012.11079859.
- [6] J.-E. Shin and J. K. Kim, "How a good sleep predicts life satisfaction: The role of zero-sum beliefs about happiness," *Frontiers Psychol.*, vol. 9, p. 1589, Aug. 2018, doi: 10.3389/fpsyg.2018.01589.
- [7] T. Daniela, C. Alessandro, C. Giuseppe, M. Fabio, M. Cristina, D. G. Luigi, and F. Michele, "Lack of sleep affects the evaluation of emotional stimuli," *Brain Res. Bull.*, vol. 82, nos. 1–2, pp. 104–108, Apr. 2010, doi: 10.1016/j.brainresbull.2010.01.014.
- [8] C. M. Barnes, J. Schaubroeck, M. Huth, and S. Ghumman, "Lack of sleep and unethical conduct," *Organizational Behav. Hum. Decis. Processes*, vol. 115, no. 2, pp. 169–180, Jul. 2011, doi: 10.1016/j.obhdp.2011.01.009.
- [9] F. Akhtar, P. K. Patel, M. B. B. Heyat, S. Yousaf, A. A. Baig, R. A. Mohona, M. M. Mutoffar, T. Bhattacharya, B. N. Teelhawod, J. P. Li, M. A. Kamal, and K. Wu, "Smartphone addiction among students and its harmful effects on mental health, oxidative stress, and neurodegeneration towards future modulation of anti-addiction therapies: A comprehensive survey based on SLR, research questions, and network visualization techniques," *CNS Neurological Disorders Drug Targets*, vol. 21, Jun. 2022, doi: 10.2174/1871527321666220 614121439.
- [10] O. AlShorman, M. Masadeh, M. B. B. Heyat, F. Akhtar, H. Almahasneh, G. M. Ashraf, and A. Alexiou, "Frontal lobe real-time EEG analysis using machine learning techniques for mental stress detection," *J. Integrative Neurosci.*, vol. 21, no. 1, p. 20, Jan. 2022, doi: 10.31083/j.jin 2101020.
- [11] F. Akhtar, M. B. Bin Heyat, J. P. Li, P. K. Patel, Rishipal, and B. Guragai, "Role of machine learning in human stress: A review," in *Proc. 17th Int. Comput. Conf. Wavelet Act. Media Technol. Inf. Process. (ICCWAMTIP)*, Dec. 2020, pp. 170–174, doi: 10.1109/ICCWAMTIP51612.2020. 9317396.
- [12] O. AlShorman, M. Masadeh, A. Alzyoud, M. B. Bin Heyat, F. Akhtar, and Rishipal, "The effects of emotional stress on learning and memory cognitive functions: An EEG review study in education," in *Proc. 6th Int. Conf. e-Learn. (econf)*, Dec. 2020, pp. 177–182, doi: 10.1109/econf51404.2020.9385468.
- [13] C. Baglioni, W. Regen, A. Teghen, K. Spiegelhalder, B. Feige, C. Nissen, and D. Riemann, "Sleep changes in the disorder of insomnia: A metaanalysis of polysomnographic studies," *Sleep Med. Rev.*, vol. 18, no. 3, pp. 195–213, Jun. 2014, doi: 10.1016/j.smrv.2013.04.001.
- [14] M. B. B. Heyat, Insomnia: Medical Sleep Disorder & Diagnosis, 1st ed. Hamburg, Germany: Anchor Academic Publishing, 2016.
- [15] B. B. Heyat, F. Akhtar, S. K. Singh, and M. M. Siddiqui, "Hamming Window are used in the Prognostic of Insomnia," in *Proc. Int. Seminar Present Scenario Future Prospectives Res. Eng. Sci. (ISPSFPRES)*, 2017, pp. 65–71.
- [16] M. B. Bin Heyat, F. Akhtar, M. Sikandar, H. Siddiqui, and S. Azad, "An overview of Dalk therapy and treatment of insomnia by Dalk therapy," Shenzhen Univ., Shenzhen, China, Tech. Rep., 2015.
- [17] S. Chung, N. I. Bohnen, R. L. Albin, K. A. Frey, M. L. T. M. Müller, and R. D. Chervin, "Insomnia and sleepiness in Parkinson disease: Associations with symptoms and comorbidities," *J. Clin. Sleep Med.*, vol. 9, no. 11, pp. 1131–1137, Nov. 2013, doi: 10.5664/jcsm.3150.
- [18] C. M. Morin, C. L. Drake, A. G. Harvey, A. D. Krystal, R. Manber, D. Riemann, and K. Spiegelhalder, "Insomnia disorder," *Nature Rev. Disease Primers*, vol. 1, no. 1, pp. 1–18, Dec. 2015, doi: 10.1038/nrdp.2015.26.
- [19] L. Scrima and T. J. Swick, "Narcolepsy," in *Synopsis of Sleep Medicine*. Poznan, Poland: Poznan Univ. of Medical Sciences, 2016, pp. 279–312, doi: 10.28942/nnj.v1i2.236.
- [20] M. M. Ohayon and C. F. Reynolds, "Epidemiological and clinical relevance of insomnia diagnosis algorithms according to the DSM-IV and the international classification of sleep disorders (ICSD)," *Sleep Med.*, vol. 10, no. 9, pp. 952–960, Oct. 2009, doi: 10.1016/j.sleep.2009.07.008.
- [21] C. M. Morin, G. Belleville, L. Belanger, and H. Ivers, "The insomnia severity index: Psychometric indicators to detect insomnia cases and evaluate treatment response," *Sleep*, vol. 34, no. 5, pp. 601–608, May 2011, doi: 10.1093/sleep/34.5.601.
- [22] S. Aydın, H. M. Saraoğlu, and S. Kara, "Singular spectrum analysis of sleep EEG in insomnia," *J. Med. Syst.*, vol. 35, no. 4, pp. 457–461, Aug. 2011, doi: 10.1007/s10916-009-9381-7.

- [23] B. Israel, D. J. Buysse, R. T. Krafty, A. Begley, J. Miewald, and M. Hall, "Short-term stability of sleep and heart rate variability in good sleepers and patients with insomnia: For some measures, one night is enough," *Sleep*, vol. 35, no. 9, pp. 1285–1291, Sep. 2012, doi: 10.5665/sleep.2088.
- [24] O. M. Al Shorman and A. M. Al shorman, "Frontal lobe and long-term memory retrieval analysis during pre-learning stress using EEG signals," *Bull. Electr. Eng. Informat.*, vol. 9, no. 1, pp. 141–145, Feb. 2020, doi: 10.11591/eei.v9i1.1335.
- [25] A. Bhattacharjee, S. Saha, S. A. Fattah, W.-P. Zhu, and M. O. Ahmad, "Sleep apnea detection based on rician modeling of feature variation in multiband EEG signal," *IEEE J. Biomed. Health Informat.*, vol. 23, no. 3, pp. 1066–1074, May 2019, doi: 10.1109/JBHI.2018.2845303.
- [26] O. AlShorman, T. Ali, and M. Irfan, "EEG analysis for pre-learning stress in the brain," in *Communications in Computer and Information Science*, vol. 752. Islamabad, Pakistan: COMSATS Univ. Islamabad, 2017, pp. 447–455, doi: 10.1007/978-981-10-6502-6_39.
- [27] M. M. Siddiqui, G. Srivastava, and S. H. Saeed, "Diagnosis of insomnia sleep disorder using short time frequency analysis of PSD approach applied on EEG signal using channel ROC-LOC," *Sleep Sci.*, vol. 9, no. 3, pp. 186–191, 2016, doi: 10.1016/j.slsci.2016.07.002.
- [28] A. Gemignani, M. Laurino, F. Provini, A. Piarulli, G. Barletta, P. d'Ascanio, R. Bedini, R. Lodi, D. N. Manners, P. Allegrini, D. Menicucci, and P. Cortelli, "Thalamic contribution to sleep slow oscillation features in humans: A single case cross sectional EEG study in fatal familial insomnia," *Sleep Med.*, vol. 13, no. 7, pp. 946–952, Aug. 2012, doi: 10.1016/j.sleep.2012.03.007.
- [29] R. F. Kaplan, Y. Wang, K. A. Loparo, M. R. Kelly, and R. R. Bootzin, "Performance evaluation of an automated single-channel sleep-wake detection algorithm," *Nature Sci. Sleep*, vol. 6, pp. 113–122, Oct. 2014, doi: 10.2147/NSS.S71159.
- [30] S. T.-B. Hamida, B. Ahmed, D. Cvetkovic, E. Jovanov, G. Kennedy, and T. Penzel, "A new era in sleep monitoring: The application of mobile technologies in insomnia diagnosis," Texas A&M Univ., Doha, Qatar, Tech. Rep., 2015, pp. 101–127.
- [31] S. T.-B. Hamida, B. Ahmed, and T. Penzel, "A novel insomnia identification method based on Hjorth parameters," in *Proc. IEEE Int. Symp. Signal Process. Inf. Technol. (ISSPIT)*, Dec. 2015, pp. 548–552, doi: 10.1109/ISSPIT.2015.7394397.
- [32] M. Bahrami and M. Forouzanfar, "Sleep apnea detection from single-lead ECG: A comprehensive analysis of machine learning and deep learning algorithms," *IEEE Trans. Instrum. Meas.*, vol. 71, pp. 1–11, 2022, doi: 10.1109/TIM.2022.3151947.
- [33] Y. Kaya, F. Kuncan, and R. Tekin, "A new approach for congestive heart failure and arrhythmia classification using angle transformation with LSTM," *Arabian J. Sci. Eng.*, vol. 47, no. 8, pp. 10497–10513, Aug. 2022, doi: 10.1007/s13369-022-06617-8.
- [34] N. Salari, A. Hosseinian-Far, M. Mohammadi, H. Ghasemi, H. Khazaie, A. Daneshkhah, and A. Ahmadi, "Detection of sleep apnea using machine learning algorithms based on ECG signals: A comprehensive systematic review," *Expert Syst. Appl.*, vol. 187, Jan. 2022, Art. no. 115950, doi: 10.1016/j.eswa.2021.115950.
- [35] N. Demir, M. Kuncan, Y. Kaya, and F. Kuncan, "Multi-layer co-occurrence matrices for person identification from ECG signals," *Traitement du Signal*, vol. 39, no. 2, pp. 431–440, Apr. 2022, doi: 10.18280/ts.390204.
- [36] D. Lai, X. Zhang, Y. Zhang, and M. B. Bin Heyat, "Convolutional neural network based detection of atrial fibrillation combing R-R intervals and F-wave frequency spectrum," in *Proc. 41st Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. (EMBC)*, Jul. 2019, pp. 4897–4900, doi: 10.1109/EMBC.2019.8856342.
- [37] D. Lai, "Prognosis of sleep bruxism using power spectral density approach applied on EEG signal of both EMG1-EMG2 and ECG1-ECG2 channels," *IEEE Access*, vol. 7, pp. 82553–82562, 2019, doi: 10.1109/ACCESS.2019.2924181.
- [38] M. B. Bin Heyat, D. Lai, F. Akhtar, M. A. Bin Hayat, and S. Azad, "Short time frequency analysis of theta activity for the diagnosis of bruxism on EEG sleep record," in *Studies in Computational Intelligence*, vol. 875, K. D. Gupta and A. Hassanien, Ed. Shenzhen, China: Shenzhen Univ., 2020, pp. 63–83.
- [39] A. L. Goldberger, L. A. N. Amaral, L. Glass, J. M. Hausdorff, P. C. Ivanov, R. G. Mark, J. E. Mietus, G. B. Moody, C.-K. Peng, and H. E. Stanley, "PhysioBank, PhysioToolkit, and PhysioNet: Components of a new research resource for complex physiologic signals," *Circulation*, vol. 101, no. 23, pp. e215–e220, Jun. 2000, doi: 10.1161/01.cir.101.23.e215.

- [40] H. Bb, F. Akhtar, A. Mehdi, S. Azad, S. Azad, and S. Azad, "Normalized power are used in the diagnosis of insomnia medical sleep syndrome through EMG1-EMG2 channel," *Austin J. Sleep Disorders*, vol. 4, no. 1, pp. 2–4, 2017.
- [41] M. B. Bin Heyat, F. Akhtar, M. Ammar, B. Hayat, and S. Azad, "Power spectral density are used in the investigation of insomnia neurological disorder," in *Proc. Congr. Symp.*, 2016, pp. 45–50.
- [42] H. Ullah, M. B. B. Heyat, H. Al Salman, H. M. Khan, F. Akhtar, A. Gumaei, A. Mehdi, A. Y. Muaad, M. S. Islam, A. Ali, Y. Bu, D. Khan, T. Pan, M. Gao, Y. Lin, and D. Lai, "An effective and lightweight deep electrocardiography arrhythmia recognition model using novel special and native structural regularization techniques on cardiac signal," *J. Healthcare Eng.*, vol. 2022, pp. 1–18, Apr. 2022, doi: 10.1155/2022/3408501.
- [43] K. L. Dodds, C. B. Miller, S. D. Kyle, N. S. Marshall, and C. J. Gordon, "Heart rate variability in insomnia patients: A critical review of the literature," *Sleep Med. Rev.*, vol. 33, pp. 88–100, Jun. 2017, doi: 10.1016/j.smrv.2016.06.004.
- [44] M. B. B. Heyat, F. Akhtar, M. A. Ansari, A. Khan, F. Alkahtani, H. Khan, and D. Lai, "Progress in detection of insomnia sleep disorder: A comprehensive review," *Current Drug Targets*, vol. 22, no. 6, pp. 672–684, Apr. 2021, doi: 10.2174/1389450121666201027125828.
- [45] D. Lai, Y. Zhang, X. Zhang, Y. Su, and M. B. Bin Heyat, "An automated strategy for early risk identification of sudden cardiac death by using machine learning approach on measurable arrhythmic risk markers," *IEEE Access*, vol. 7, pp. 94701–94716, 2019, doi: 10.1109/ACCESS.2019.2925847.
- [46] M. B. B. Heyat, F. Akhtar, M. H. Khan, N. Ullah, I. Gul, H. Khan, and D. Lai, "Detection, treatment planning, and genetic predisposition of bruxism: A systematic mapping process and network visualization technique," *CNS Neurol. Disorders Drug Targets*, vol. 20, no. 8, pp. 755–775, Oct. 2021, doi: 10.2174/1871527319666201110124954.
- [47] M. B. B. Heyat, F. Akhtar, S. J. Abbas, M. Al-Sarem, A. Alqarafi, A. Stalin, R. Abbasi, A. Y. Muaad, D. Lai, and K. Wu, "Wearable flexible electronics based cardiac electrode for researcher mental stress detection system using machine learning models on single lead electrocardiogram signal," *Biosensors*, vol. 12, no. 6, p. 427, Jun. 2022, doi: 10.3390/bios12060427.
- [48] P. D. Welch, "Welch_1967_modified_periodogram_method," Trans. Audio Electroacoust., IBM Res. Center, Yorktown Heights, NY, USA, Tech. Rep., 1967.
- [49] M. B. B. Heyat, D. Lai, F. Akhtar, M. A. B. Hayat, S. Azad, S. Azad, and S. Azad, "Bruxism detection using single-channel C4-A1 on human sleep S2 stage recording," in *Intelligent Data Analysis*, D. Gupta, S. Bhattacharyya, and A. Khanna, Eds., 1st ed. Hoboken, NJ, USA: Wiley, 2020, pp. 347–367.
- [50] M. B. B. Heyat, F. Akhtar, and S. Azad, "Comparative analysis of original wave and filtered wave of EEG signal used in the prognostic of bruxism medical sleep syndrome," *Int. J. Trend Sci. Res. Develop.*, vol. 1, no. 1, pp. 7–9, Dec. 2016, doi: 10.31142/ijtsrd53.
- [51] M. B. B. Heyat, D. Lai, F. I. Khan, and Y. Zhang, "Sleep bruxism detection using decision tree method by the combination of C4-P4 and C4-A1 channels of scalp EEG," *IEEE Access*, vol. 7, pp. 102542–102553, 2019, doi: 10.1109/ACCESS.2019.2928020.
- [52] L. Breiman, "Random forests," Mach. Learn., vol. 45, no. 1, pp. 5–32, 2001, doi: 10.1023/A:1010933404324.
- [53] T. K. Ho, "The random subspace method for constructing decision forests," *IEEE Trans. Pattern Anal. Mach. Intell.*, vol. 20, no. 8, pp. 832–844, Aug. 1998, doi: 10.1109/34.709601.
- [54] T. K. Ho, "A data complexity analysis of comparative advantages of decision forest constructors," *Pattern Anal. Appl.*, vol. 5, no. 2, pp. 102–112, 2002, doi: 10.1007/s100440200009.
- [55] D. Ramage, D. Hall, R. Nallapati, and C. D. Manning, "Labeled LDA," Stanford Univ., Stanford, CA, USA, Tech. Rep., 2009, p. 248, doi: 10.3115/1699510.1699543.
- [56] A. K. Nawabi, S. Jinfang, R. Abbasi, M. S. Iqbal, M. B. B. Heyat, F. Akhtar, K. Wu, and B. A. Twumasi, "Segmentation of drug-treated cell image and mitochondrial-oxidative stress using deep convolutional neural network," *Oxidative Med. Cellular Longevity*, vol. 2022, pp. 1–14, May 2022, doi: 10.1155/2022/5641727.
- [57] M. S. Iqbal, R. Abbasi, M. B. B. Heyat, F. Akhtar, A. S. Abdelgeliel, S. Albogami, and E. Fayad, and Muhammad Atif Iqbal, "Recognition of mRNA N4 acetylcytidine (AC4C) by using non-deep vs. deep learning," *Appl. Sci.*, vol. 12, no. 3, pp. 1–16, 2022, doi: 10.3390/app12031344.

- [58] R. Pal, M. B. B. Heyat, Z. You, B. Pardhan, F. Akhtar, S. J. Abbas, B. Guragai, and K. Acharya, "Effect of maha mrityunjaya HYMN recitation on human brain for the analysis of single EEG channel C4-A1 using machine learning classifiers on yoga practitioner," in *Proc. 17th Int. Comput. Conf. Wavelet Act. Media Technol. Inf. Process. (ICCWAMTIP)*, Dec. 2020, pp. 89–92, doi: 10.1109/ICCWAMTIP51612.2020.9317384.
- [59] A. Sultana, K. Rahman, M. B. B. Heyat, F. Akhtar, and A. Y. Muaad, "Role of inflammation, oxidative stress, and mitochondrial changes in premenstrual psychosomatic behavioral symptoms with antiinflammatory, antioxidant herbs, and nutritional supplements," *Oxidative Med. Cellular Longevity*, vol. 2022, Jul. 2022, Art. no. 3599246, doi: 10.1155/2022/3599246.
- [60] A. Sultana, W. Begum, R. Saeedi, K. Rahman, M. B. B. Heyat, F. Akhtar, N. T. Son, and H. Ullah, "Experimental and computational approaches for the classification and correlation of temperament (Mizaj) and uterine dystemperament (Su'-I-Mizaj Al-Rahim) in abnormal vaginal discharge (Sayalan Al-Rahim) based on clinical analysis using support vector machine," *Complexity*, vol. 2022, pp. 1–16, Aug. 2022, doi: 10.1155/2022/5718501.
- [61] C. C. Ukwuoma, Q. Zhiguang, M. B. Bin Heyat, L. Ali, Z. Almaspoor, and H. N. Monday, "Recent advancements in fruit detection and classification using deep learning techniques," *Math. Problems Eng.*, vol. 2022, pp. 1–29, Jan. 2022, doi: 10.1155/2022/9210947.
- [62] R. Mehrrotraa, M. A. Ansari, R. Agrawal, P. Tripathi, M. B. B. Heyat, M. Al-Sarem, A. Y. M. Muaad, W. A. E. Nagmeldin, A. Abdelmaboud, and F. Saeed, "Ensembling of efficient deep convolutional networks and machine learning algorithms for resource effective detection of tuberculosis using thoracic (chest) radiography," *IEEE Access*, vol. 10, pp. 85442–85458, 2022, doi: 10.1109/ACCESS.2022.3194152.
- [63] A. R. Hassan and M. I. H. Bhuiyan, "Computer-aided sleep staging using complete ensemble empirical mode decomposition with adaptive noise and bootstrap aggregating," *Biomed. Signal Process. Control*, vol. 24, pp. 1–10, Feb. 2016, doi: 10.1016/j.bspc.2015.09.002.
- [64] M. Schrader, C. Zywietz, V. Von Einem, B. Widiger, and G. Joseph, "Detection of sleep apnea in single channel ECGs from the physionet data base," in *Computers in Cardiology*. Örebro, Sweden: Örebro Univ., 2000, pp. 263–266, doi: 10.1109/cic.2000.898507.
- [65] S. A. Imtiaz and E. Rodriguez-Villegas. (Dec. 2015). Low-Complexity Algorithms for Automatic Detection of Sleep Stages and Events for use in Wearable EEG Systems. [Online]. Available: https://spiral.imperial.ac.uk/bitstream/10044/1/29459/1/Imtiaz-SA-2016-PhD-Thesis.pdf
- [66] M. Längkvist, L. Karlsson, and A. Loutfi, "Sleep stage classification using unsupervised feature learning," *Adv. Artif. Neural Syst.*, vol. 2012, pp. 1–9, Jan. 2012, doi: 10.1155/2012/107046.
- [67] A. J. Boe, L. L. McGee Koch, M. K. O'Brien, N. Shawen, J. A. Rogers, R. L. Lieber, K. J. Reid, P. C. Zee, and A. Jayaraman, "Automating sleep stage classification using wireless, wearable sensors," *NPJ Digit. Med.*, vol. 2, no. 1, pp. 1–9, Dec. 2019, doi: 10.1038/s41746-019-0210-1.
- [68] V. Bajaj and R. B. Pachori, "Automatic classification of sleep stages based on the time-frequency image of EEG signals," *Comput. Meth*ods Programs Biomed., vol. 112, no. 3, pp. 320–328, 2013, doi: 10.1016/j.cmpb.2013.07.006.
- [69] Y. Mitsukura, K. Fukunaga, M. Yasui, and M. Mimura, "Sleep stage detection using only heart rate," *Health Informat. J.*, vol. 26, no. 1, pp. 376–387, Mar. 2020, doi: 10.1177/1460458219827349.
- [70] H. Abdullah, T. Penzel, and D. Cvetkovic, "Detection of insomnia from EEG and ECG," in *Proc. IFMBE*, vol. 43, 2014, pp. 687–690, doi: 10.1007/978-3-319-02913-9_175.
- [71] M. Shahin, B. Ahmed, S. T.-B. Hamida, F. L. Mulaffer, M. Glos, and T. Penzel, "Deep learning and insomnia: Assisting clinicians with their diagnosis," *IEEE J. Biomed. Health Informat.*, vol. 21, no. 6, pp. 1546–1553, Nov. 2017, doi: 10.1109/JBHI.2017.2650199.
- [72] A. R. Hassan and M. I. H. Bhuiyan, "An automated method for sleep staging from EEG signals using normal inverse Gaussian parameters and adaptive boosting," *Neurocomputing*, vol. 219, pp. 76–87, Jan. 2017, doi: 10.1016/j.neucom.2016.09.011.
- [73] J. Zhang, R. Yao, W. Ge, and J. Gao, "Orthogonal convolutional neural networks for automatic sleep stage classification based on singlechannel EEG," *Comput. Methods Programs Biomed.*, vol. 183, Jan. 2020, Art. no. 105089, doi: 10.1016/j.cmpb.2019.105089.
- [74] J. Zhou, G. Wang, J. Liu, D. Wu, W. Xu, Z. Wang, J. Ye, M. Xia, Y. Hu, and Y. Tian, "Automatic sleep stage classification with single channel EEG signal based on two-layer stacked ensemble model," *IEEE Access*, vol. 8, pp. 57283–57297, 2020, doi: 10.1109/ACCESS.2020.2982434.



PRAGATI TRIPATHI received the M.Tech. degree in power electronics from Gautam Buddha University, in 2018. She is currently working as a Research Scholar with the School of Engineering, Gautam Buddha University, Gr. Noida. She has been an Associate with IIT Delhi and served as a Research Associate at Sharda University Greater Noida, India. Her research interests include signal processing, brain mapping, and neuroscience.



M. A. ANSARI (Senior Member, IEEE) received the B.Tech. degree in electrical engineering from A.M.U. Aligarh India, and the M.Tech. and Ph.D. degrees in medical image processing from the Indian Institute of Technology, Roorkee, India. He is currently an Associate with the School of Engineering, Gautam Buddha University, Gr. Noida, NCR Delhi, India. He has more than 20 years of teaching experience in engineering and technology. He has published more

than 200 research papers in reputed national and international journals/conferences and several patents. His research interests include digital image processing, medical imaging, biomedical instrumentation, AI, ML, soft computing, and wavelet applications to engineering problems. He is a senior member of ISIAM and the HAC/SIGHT Chair in Uttar Pradesh, India.



TAPAN KUMAR GANDHI received the Ph.D. degree in biomedical engineering from IIT Delhi. He is currently working as an Associate Professor with the Department of Electrical Engineering, IIT Delhi, and also an Adjunct Faculty with the School of Information Technology, IIT Delhi, as well as Adjunct Professor, IIT Jodhpur. Following the Ph.D., he has spent more than three years as a Scientist at MIT, USA. He received the Ph.D. Fellowship from MIT. He was awarded the INSPIRE

Faculty in the engineering and technology category of the Department of Science and Technology, Government of India. During this five years tenure, he was awarded as the Excellent INSPIRE Faculty by DST, Government of India.



RAJAT MEHROTRA received the B.Tech. degree in electrical and electronics engineering and the M.Tech. degree in telecommunication engineering from the Dr. A.P.J. Abdul Kalam Technical University, Lucknow (Formerly UPTU), in 2008 and 2014, respectively. He is currently an Assistant Professor with the Electrical and Electronics Engineering Department, GL Bajaj Institute of Technology and Management, Greater Noida, India. His research interests include digital image

processing, biomedical imaging, and deep learning. He is currently involved in research with the School of Engineering, Gautam Buddha University, Greater Noida.



MD BELAL BIN HEYAT received the B.Tech. degree in electronics and instrumentation (EI) and the M.Tech. degree in electronics circuit and system (ECS) from Integral University, Lucknow, Uttar Pradesh, India, in 2014 and 2016, respectively, and the Ph.D. degree in electronic science and technology from the University of Electronic Science and Technology of China, Chengdu, Sichuan, China.

He is currently working as a Postdoctoral Researcher with the College of Computer Science and Software Engineering,

Shenzhen University, China. He is also working as a Visiting Postdoctoral Researcher at IIIT Hyderabad, India, and as a Faculty Member with the Department of Science and Engineering, Novel Global Community Educational Foundation, Hebersham, NSW, Australia. He has published more than forty articles in reputed international journals and conferences. He is engaged in the areas of detection, sleep disorders, neurological disorders, psycho-neurological disorders, cardiovascular diseases, signal processing, and medical machine learning. He received nine awards and also worked as a Research Associate, the Vice-Chairman of the UESTC Country League, and a Country Representative of India during the Ph.D. at the University of Electronic Science and Technology of China. He has been serving as a Guest Editor/Reviewer for over ten journals including Applied Sciences, Biosensors, Cells, Journal of Integrative Neuroscience, Electronics, Computers in Biology and Medicine, IEEE Access, Sensors, Reviews in Cardiovascular Medicine, International Journal of Environmental Research and Public Health, Journal of Healthcare Engineering, and AI.



FAIJAN AKHTAR received the Bachelor of Technology degree in information and technology from Integral University, Lucknow, Uttar Pradesh, India, and the Master of Technology degree in computer and science engineering from Jamia Hamdard University, New Delhi, India. He is currently pursuing the Ph.D. degree with the University of Electronic Science and Technology of China. He has published several research papers in reputed international journals and conferences.

His current research interests include machine learning, deep learning, and blocks chain technology.



CHIAGOZIEM C. UKWUOMA received the B.Eng. degree in mechanical engineeringautomobile technology from the Federal University of Technology Owerri Nigeria and the M.Sc. degree in software engineering from the University of Electronic Science and Technology of China (UESTC), where he is currently pursuing the Ph.D. degree. His research interests include object detection and classification, attention mechanism, and medical imaging.



ABDULLAH Y. MUAAD received the B.Sc. degree in computer science from Hail University, in 2008, and the M.Sc. degree from the University of Mysore (UOM), Mysuru, Karnataka, India, in 2018, where he is currently pursuing the Ph.D. degree with the Department of Computer Science. Also, he worked as a General Registrar and a Lecturer at the Sanaa Community College, Yemen, from 2010 to 2014. Also, he is a reviewer in many SCI journals. He has authored and coauthored

more than five journal and conference papers in well-reputed international journals. His main areas of interest and current research interests include natural language processing, text representation, image processing, and machine and deep learning. He received five awards during his study.



YASSER M. KADAH (Senior Member, IEEE) received the B.Sc. and M.Sc. degrees from the Biomedical Engineering Department, Cairo University, in 1989 and 1992, respectively, and the Ph.D. degree in biomedical engineering from the University of Minnesota, in 1997. He is currently a Professor of biomedical engineering with the Department of Electrical and Computer Engineering, King Abdulaziz University, Jeddah, Saudi Arabia. He worked as a Research Assis-

tant with the Department of Radiology, University of Minnesota Medical School, from 1996 to 1997, and also a Postdoctoral Fellow at the Center of Magnetic Resonance Research, University of Minnesota, in 1998. He became an Assistant Professor of biomedical engineering at Cairo University, from 1998 to 2003. He also worked during the same period as the Director of Researcher and a Developer at IBE Technologies, Egypt. He was also a Research Associate at the Biomedical Imaging Technology Center, Biomedical Engineering Department, Emory University, and the Georgia Institute of Technology, from December 2002 to July 2004. He is/was a Professor of biomedical engineering at Cairo University for many years and the Former Department Chair. His research interests include medical imaging and in particular MRI and ultrasound imaging and multi-dimensional signal processing for biomedical applications such as brain-computer interface. He is a Senior Member of the International Society of Optical Engineering (SPIE), since 2015. He is also currently an Active Member of the Association for the Advancement of Medical Instrumentation (AAMI). Along his career, he received several awards and recognitions including the record for highest undergraduate GPA with the Department of Biomedical Engineering, Cairo University, the Islamic Development Bank Merit Scholarship (1993-1996), his biography was selected to appear in Marquis Who's Who in the World in 2002, winning the E.K. Zaviosky stipend from the international society of magnetic resonance in imaging, in 2002, and awarded the 2005 National Encouragement Award from the Academy of Scientific Research and Technology, Ministry of Higher Education and Research, Egypt, as well as the Award for Best Professor in biomedical engineering from the Cairo University Faculty Club, in 2007. He was also selected to receive the 2008 Scientific Distinction Award from Cairo University.



MUGAHED A. AL-ANTARI received the Ph.D. degree from the Biomedical Engineering Department, College of Electronics and Information, Kyung Hee University (KHU), Republic of Korea, in August 2019. He is currently an Assistant Professor with the Department of Artificial Intelligence (AI), College of Software and Convergence Technology, Sejong University, Seoul, Republic of Korea. He has teaching experience as a Lecturer at the Department of Biomedical Engineering with

different public and private universities for more than five years. Meanwhile, he has professional industry experience as a Senior Researcher and the Executive Engineering Director for more than five years for developing mini- and whole-body DXA machines using X-ray pencil and fan beams at the YOZMA BMTech Group, Seoul. He joined the Ubiquitous Computing Laboratory (UC Lab) as a Postdoctoral Researcher at the Department of Computer Science and Engineering, Kyung Hee University. His recent AI-based publications have earned a lot of attention from the international journal editorials and has been selected to represent a cornerstone for the modern medicine. His current research interests include explainable artificial intelligence (XAI), deep reinforcement learning, deep learning, machine learning, pattern recognition, medical signal and image processing, and medical imaging of dual-energy X-ray absorptiometry (DXA). He is a member of IEEE EMBS, since 20214. In August 2019, he received the Ph.D. degree from an excellent Thesis Award from Biomedical Engineering Department, College of Electronics and Information, Kyung Hee University (KHU).



JIAN PING LI was born in October 1964. He received the B.S. degree in applied mathematics from Chongqing University, Chongqing, China, in 1986, the M.Eng. degree in software engineering and the M.Sc. degree in computational mathematics from the Graduate School of Xi'an Jiaotong University, Xi'an, Shanxi, China, in 1989, and the Ph.D. degree in computer science from the Graduate School of Chongqing University, Chongqing, in 1998. From 1999 to 2000,

he was a Postdoctoral Researcher and a Visiting Professor at Hong Kong Baptist University, under the leadership of Professor Dr. Yuan Yan Tang. From 2002-2015, he was a Senior Visiting Professor at Provence University, France; the University of Guelph, Canada, Washington University, USA; and the University of Zurich, Switzerland. He is currently a Professor with the International Centre for Wavelets Analysis and its Applications (ICWAA), University of Electronic Science and Technology of China (UESTC), Chengdu, China, and as a Head of ICWAA, he is the Dean of Big Data Research Institute, UESTC. His book Wavelet Analysis and Signal Processing: Theory, Applications and Software Implementations, was awarded the second prize of the National Science and Technology of China, in 1999, and it is very important reference for many scientists. He has published more than 300 technical papers and is the author/coauthor of 30 books on subjects ranging from wavelet analysis and its applications to computer science. His current interests include wavelet theory and applications, fractal, image processing, pattern recognition, information security, big data, and artificial intelligence. He has served as a Syndic for Awards of National Science and Technology of China, a Founder of the International Journal on Multiresolution, Wavelets, Applications (IJMWA), a Founder and an Editorin-Chief of International Progress on Wavelet Active Media Technology and Information Processing, and also as the general chairman of many international conferences. He was a reviewer for many international famous Journals.

...