

# Advances in Computational Human Phantoms and Their Applications in Biomedical Engineering—A Topical Review

Wolfgang Kainz<sup>1</sup>, Member, IEEE, Esra Neufeld, Member, IEEE, Wesley E. Bolch, Member, IEEE, Christian G. Graff, Chan Hyeong Kim, Member, IEEE, Niels Kuster, Fellow, IEEE, Bryn Lloyd, Member, IEEE, Tina Morrison, Member, IEEE, Paul Segars, Member, IEEE, Yeon Soo Yeom, Member, IEEE, Maria Zankl, Member, IEEE, X. George Xu, and Benjamin M. W. Tsui, Fellow, IEEE

**Abstract**—Over the past decades, significant improvements have been made in the field of computational human phantoms (CHPs) and their applications in biomedical engineering. Their sophistication has dramatically increased. The very first CHPs were composed of simple geometric volumes, e.g., cylinders and spheres, while current CHPs have a high resolution, cover a substantial range of the patient population, have high anatomical accuracy, are poseable, morphable, and are augmented with various details to perform functionalized computations. Advances in imaging techniques and semiautomated segmentation tools allow fast and personalized development of CHPs. These advances open the door to quickly develop personalized CHPs, inherently including the disease of the patient. Because many of these CHPs are increasingly providing data for regulatory submissions of various medical devices, the validity, anatomical accuracy, and availability to cover the entire patient population is of utmost importance. This paper is organized into

two main sections: the first section reviews the different modeling techniques used to create CHPs, whereas the second section discusses various applications of CHPs in biomedical engineering. Each topic gives an overview, a brief history, recent developments, and an outlook into the future.

**Index Terms**—Computational modeling, human anatomy, phantoms.

## I. MODELING TECHNIQUES FOR REALISTIC COMPUTATIONAL HUMAN PHANTOMS AS APPLIED IN BIOMEDICAL ENGINEERING

OVER the last 50 years, many computational human phantoms (CHPs) have been developed across different research communities for a variety of applications. The following section discusses various methods to create CHPs, the verification, validation, and quality assurance (QA) techniques involved with these phantoms, the modeling of tissue properties, and a survey of libraries of models currently available.

### A. Advanced Phantom Geometries

The construction of a CHP must consider multiple factors such as anatomy, tissue properties, computational efficiency, and geometrical compatibility with simulation codes, e.g., Monte Carlo (MC) or analytical. As a first step, a phantom must be generated by explicitly defining the surfaces and tissue properties of the necessary tissues in which interactions occur. For defining the tissue surfaces two general modeling methods from the computer graphics community have been widely used: 1) constructive solid geometry (CSG) and 2) boundary representation (BREP) [1]–[3]. The following sections review these two established methods and three additional evolving methods.

*1) Constructive Solid Geometry (Mathematical and Voxelized) Phantoms:* CSG allows the modeler to create a solid geometric object utilizing simple shapes, called primitives. Examples of these primitives include cuboids, cylinders, prisms, pyramids, spheres, cones, and ellipsoids—surfaces that are easily described by quadratic equations. Overlap, cut planes, and intersections of the geometric objects can be used to construct organs.

As an example, the left lung can be represented in the CSG method by “half an ellipsoid with a section removed” [4].

Manuscript received August 1, 2018; revised October 16, 2018; accepted November 11, 2018. Date of publication December 3, 2018; date of current version December 31, 2018. (Corresponding author: Wolfgang Kainz.)

W. Kainz, C. G. Graff, and T. Morrison are with the Food and Drug Administration, Center for Devices and Radiological Health, Silver Spring, MD 20993 USA (e-mail: wolfgang.kainz@fda.hhs.gov; christian.graff@fda.hhs.gov; tina.morrison@fda.hhs.gov).

E. Neufeld and B. Lloyd are with the Foundation for Research on Information Technologies in Society, 8092 Zürich, Switzerland (e-mail: neufeld@itis.ethz.ch; lloyd@itis.ethz.ch).

W. E. Bolch is with the Department of Biomedical Engineering, University of Florida, Gainesville, FL 32611 USA (e-mail: wbolch@ufl.edu).

C. H. Kim is with the Department of Nuclear Engineering, Hanyang University, Seoul 04763, South Korea (e-mail: chkim@hanyang.ac.kr).

N. Kuster is with the Department of Information Technology and Electrical Engineering, Swiss Federal Institute of Technology, 8092 Zürich, Switzerland, and the Foundation for Research on Information Technologies in Society, 8092 Zürich, Switzerland (e-mail: kuster@itis.ethz.ch).

P. Segars is with the Department of Radiology, Duke University, Durham, NC 27710 USA (e-mail: paul.segars@duke.edu).

Y. S. Yeom is with the Division of Cancer Epidemiology and Genetics, National Institutes of Health, Rockville, MD 20892 USA (e-mail: yeonsoo.yeom@nih.gov).

M. Zankl is with the Department of Medical Radiation Physics and Diagnostics, Helmholtz Zentrum München German Research Center for Environmental Health, 85764 Munich, Germany (e-mail: zankl@helmholtz-muenchen.de).

X. G. Xu is with the Department of Mechanical Aerospace and Nuclear Engineering, Rensselaer Polytechnic Institute, Troy, NY 12180 USA (e-mail: xug2@rpi.edu).

B. M. W. Tsui is with the School of Medicine, Johns Hopkins University, Baltimore, MD 21218 USA (e-mail: btsui@jhmi.edu).

Color versions of one or more of the figures in this paper are available online at <http://ieeexplore.ieee.org>.

Digital Object Identifier 10.1109/TRPMS.2018.2883437

The cut-out section can be defined by a Boolean operation subtracting one ellipsoid from the other (see Fig. 1(a) [5], [6]). These surface equations used in the CSG method are computationally efficient and compatible by nearly all MC radiation transport simulation codes. Even with complicated and carefully designed Boolean operations, phantoms based on quadratic equations have limited anatomical accuracy.

Voxels have also been used as a CSG modeling technique. Fig. 1(b) shows an example where the left lung is defined as an assembly of 3-D cuboids or voxels. Medical image data, e.g., computed tomography (CT) images, can be converted to a voxel-based geometry that provides a direct way of realistically describing the human anatomy [5], [6]. This can be done by segmenting imaging data and then assigning standard tissue properties to the segmented objects (see Section I-B3). Automatic methods have also been investigated to directly convert image values (CT numbers for example) into tissue properties for a voxel-based phantom [7]. The geometry of a voxel is very easy for existing simulation codes to handle. However, the boundary of a voxel object is defined by a stair-stepped surface instead of a smooth surface, as shown in Fig. 1(b). Thus, the anatomical fidelity depends on the voxel size, which can be an issue especially for thin and small tissues, e.g., the micrometer-thick radiosensitive target layers of the alimentary and respiratory tract organs. In addition, an adjustment to the organ shape will likely involve all underlying voxels, which is computationally inefficient.

2) *Boundary Representation (NURBS or Mesh) Phantoms:* The human anatomy can also be defined by advanced BREP modeling techniques involving nonuniform rational b-spline (NURBS) or polygon mesh surfaces. The most common technique to create a BREP-based phantom involves the surface contour extraction of each organ from a tomographic image dataset. NURBS are lofted through the contours defined for each structure creating smooth, continuous surfaces. The individual surfaces for the tissues are then integrated into a whole-body assembly. In essence, the contours convert the voxels into NURBS that are smooth and anatomically realistic. These phantoms are commonly referred to as “NURBS,” “mesh,” or “BREP” phantoms. Fig. 1(c) shows the triangular mesh of a left lung, derived from high-resolution tomographic images [5], [6].

In BREP-based design, the exterior of an object can be defined as NURBS, which offer very smooth surfaces. These surfaces can alternatively be represented as polygons whose vertices are defined by a set of coordinate values  $x$ ,  $y$ , and  $z$ . In principle, NURBS and polygon meshes are interchangeable BREP data structures. Unlike the CSG representation, BREP is much more flexible because a richer set of operation tools are available, e.g., extrusion, chamfering, blending, drafting, shelling, and tweaking. These features allow BREP-based models to include very complex anatomical features. Furthermore, the BREP technique is ideally suited for surface deformation. NURBS and polygon meshes can be altered easily by applying transformations (rigid or nonrigid) to the surface or vertex points. Such flexibility is necessary for the adjustment of organ size, for organ motion simulations, and for changing the posture of phantoms.




Fig. 1. Model of the left lung defined by different modeling methods. (a) CSG-type modeling after the Boolean operation (subtraction) is performed involving two ellipsoids. (b) Voxel representation of the lung. (c) BREP-type of modeling of the same lung using a polygon mesh [5], [6].

3) *Morphing and Posing Techniques:* A range of applications require CHPs with different characteristics, such as posture, height, weight, body-mass-index (BMI), or organ volume and shape. Unfortunately, significant effort is required to create new whole-body models, e.g., seven years for the Visible Korean male phantom [8]. For this purpose, morphing techniques have been developed, which permit modification of the volume and shape of tissues of existing reference phantoms. Three types of morphing and posing approaches are found in the literature: 1) geometrical approaches; 2) physics-based methods; and 3) image registration mapping techniques.

Geometrical methods vary in complexity and anatomical realism, ranging from simple heuristics for scaling and repositioning of individual organs, to computer graphics deformation approaches. Na *et al.* [9] created 100 adult phantoms based on the Rensselaer Polytechnic Institute (RPI) male and female reference models. Body dimensions and individual organ volumes were deformed to match statistical volumes from the NHANES study [10] using iterative scaling along the tissue surface normal, including collision detection to avoid overlapping tissues. Similar techniques were used to create a library of pediatric models based on reference models from the virtual population (ViP) [11], [12]. Interactive tools to change the posture of anatomical models permit positioning of bones around articulated joints in real-time and interpolate the tissue deformation using techniques such as dual quaternion skinning [13]–[15].

A second category includes methods which constrain the morphing by a biomechanical tissue deformation model. In this approach, the tissue is modeled as a hyper-elastic material, with soft tissues and stiff bones. Combined with articulated joints, this formulation permits posture changes with realistic deformations. By incorporating a model of tissue growth, analogous to thermal expansion, this approach was used to change the BMI of the ViP [16].

The third group of approaches deforms a template or reference model into a target anatomy using image, or surface, registration. For example, the XCAT models are based on reference anatomies (originally the Visible Human male and female [17]) by performing image registration of the references to CT images of patients [18], [19]. Only few others have worked on personalization of full body models [20], [21]. Several approaches have been presented using mouse image data [22], [23].

4) *Organ Motion Techniques:* In medical imaging, reconstruction algorithms typically assume that everything is static




Fig. 2. (a) Cardiac and RM models of the surface-based XCAT (left) and VIP-Man (right) phantoms. Transformations are applied to the surfaces to deform them. (b) 4-D voxelized respiratory phantom of Han *et al.* [25]. Transforms are applied to the individual voxels to deform the image. Full animation of the RM and deformation can be found elsewhere ([hurel.hanyang.ac.kr/Phantom/4DVoxel.gif](http://hurel.hanyang.ac.kr/Phantom/4DVoxel.gif)).

during image acquisition. Patient movements result in image artifacts that can mimic, or mask, disease indicators. Patient motion must also be accounted for in radiation therapy. Motion of the tumor and background tissues need to be assessed in order to best target the tumor while sparing the healthy tissues. With the ability to simulate any number of motion variations, time-changing 4-D computational phantoms provide a vital research tool to investigate and understand the effects of motion, and to compensate for them.

Motion can be simulated by applying transformations (rigid and nonrigid) to the phantom’s anatomical structures. For surface-based phantoms such as the XCAT [17] and the VIP-Man [24], the transformations can be simply applied to the surface or vertex points defining the objects [see Fig. 2(a)]. For voxelized phantoms, e.g., [25], transformations are applied to each individual image voxel and interpolation is used to generate deformed images [see Fig. 2(b)].

Gated patient imaging data, such as CT or magnetic resonance imaging (MRI) has been typically used to define the time-changing transformations for the organs and structures allowing a model to realistically simulate motion. However, only one realization of the motion is captured, specifically that derived from the imaging data. To simulate variations in a given motion, parameters can be setup to alter the deformations of the surfaces, but these alterations lack a physiological basis [17].

Finite element (FE) techniques, e.g., [26] and [27], are currently being investigated to create more robust and flexible models for patient motions. FE methods simulate the physiological and physical interactions of the tissues on multiple scales from the cell to the organ. Many different parameters can be altered in a realistic, physiologically based manner, to realistically simulate normal and abnormal variations in motion.

5) *Direct Use of Mesh Phantoms in Monte Carlo Simulation:* Although providing significant advantages over the stylized and voxel phantoms, BREP-based phantoms suffer from a technical issue: the compatibility with MC particle transport codes. NURBS phantoms cannot be directly used in MC codes because implementing ray-tracing functions for particle transport in NURBS geometry is technically very challenging, requiring highly complex and time-consuming algebraic calculations. Therefore, the NURBS phantoms need to be converted to voxel phantoms, via a so-called “voxelization” process. The resulting voxel phantoms cannot preserve thin or small tissues, such as the skin target layer ( $\sim 50 \mu\text{m}$ ), the detailed eye structure, and the micrometer-thick radiosensitive target layers of the alimentary and respiratory tract organs.

The issue has been addressed at least for polygon mesh phantoms. A polygon mesh phantom, called the PSRK-Man, has been successfully implemented, without voxelization, in the Geant4 code using the G4Tessellated Solid class, which was originally designed for importing computer-aided design models [28]. The implementation was possible because the phantom is composed of perfect meshes of organs and tissues, i.e., without abnormal facets. It was then found that the polygon mesh phantom was significantly slower (e.g., 70–150 times slower for photons) in transporting particles than its counterpart voxel phantom, i.e., the HDRK-Man.

The issue of the slow computational speed was then overcome by converting the geometry format into a tetrahedral mesh (TM), and implementing the TM phantom in the Geant4 code using the G4Tet class, resulting in a 2–3 order of magnitude increase in computational speed depending on particles and energies [29]. This improvement in computational speed is mainly due to the large number of facets to be checked by the computationally expensive ray-tracing functions in the polygon mesh, reducing the TM geometry to just four facets. Acknowledging both limitations of the voxel geometry and advantages of the TM geometry, the International Commission on Radiological Protection (ICRP) recently decided to convert the voxel-type reference CHPs [30] into the TM format. The adult male and female phantoms, called mesh-type reference computational phantoms, have been completed and tested with the Geant4, MCNP6, and PHITS codes [31] for initialization time, computation speed, and memory requirements as shown in Fig. 3.

Note that the MCNP6 code, i.e., merger of MCNP5 and MCNPX, provides a new features: 1) the EMBED card, for importing an unstructured mesh geometry (e.g., TM geometry) and 2) the PHITS code providing a new option in the LATTICE card for importing the TM geometry.

The result of computational speed shows that the male TM phantom on the PHITS code was 2–3 times faster than the male voxel-type reference CHP on the PHITS code for all particles (photons, electrons, and helium ions) and energies considered. This result implies that, if ICRP keeps using the PHITS code for most dose coefficient calculations, the use of the new TM phantoms will even improve the computational speed of dose coefficients. The TM phantom on the Geant4 code was even faster than the voxel phantom on the




Fig. 3. Adult female mesh-type reference CHP implemented in Geant4 (left), MCNP6 (middle), and PHITS (right).

PHITS code for photons and electrons, by up to 20 times. However, the TM phantom on the MCNP6 code was slower than the voxel phantom on the PHITS code by up to 20 times.

The spatial variation of tissue composition and mass density in a tissue cannot be modeled in a BREP-based phantom, because BREP meshes define only the boundary of a tissue. The tissues in a BREP-based phantom are therefore all internally uniform. On the other hand, the composition and density variation in a tissue can be defined in a TM phantom using the volumetric 3-D structure of the TM [32]. The organs and tissues of the TM phantom can also be continuously moved or deformed during an MC simulation, leading to more realistic 4-D MC simulations.

### B. Numerical Aspects, Verification, and Validation

1) *Verification, Validation, and Quality Assurance:* Credibility assessment of simulations include verification and validation (V&V), among other aspects such as applicability analysis [33] and QA. Verification, simply put, asks the question, “did you implement the computational model correctly/as intended?”; whereas validation asks, “did you implement the correct computational model?” The objective of verification, as defined by the ASME V&V40 subcommittee [34], is to ensure that the computational model is implemented correctly and then accurately solved. The objective of validation is to assess the accuracy of the prediction as compared to real world experimental data. Important factors to consider include the assessment of the model form, model inputs, along with assessing the experimental comparator, which includes test conditions and test samples. The assessment of the model prediction accuracy lies in the nature of the comparison between the output from the model and the output from the experimental comparator, and the identification and quantification of the uncertainties related to both. Lastly, the relevance and “goodness” of the validation outcome is judged based on the context-of-use of the computational model, i.e., the

scope and role of the computational model for decision making or answering a question of interest, and is tied to the relevant real-world behavior of quantities-of-interest. Hence, while a computational model can be considered universally verified, CHPs can typically only be considered sufficiently validated with regard to certain given context-of-uses.

V&V of CHPs may be performed as a standalone activity when not used in simulations. However, the definitive V&V of CHPs will occur alongside the simulation and in the context of an application, mainly as part of the assessment of model form and model inputs, e.g., anatomical geometries, tissue properties, and solution verification, e.g., discretization errors. This section will discuss the concepts of V&V in the context of CHPs. Simulation V&V is outside the scope of this paper.

a) *Verification:* Verification in the context of CHPs covers the process of model generation, the involved tools, and the obtained models. An example of the first is the assessment of the image segmentation quality and processes by anatomical experts, as documented, e.g., in the QA report [35]. Examples of model generation tool verification include the following.

- 1) Assessing the fidelity of extracted surfaces to the original segmentation or the distance between initially extracted surfaces and simplified surfaces [12].
- 2) Ascertaining the fidelity and accuracy of the process of morphing an existing CHP to match specific medical image data [36], [37], e.g., in the process of building model populations or personalized models.
- 3) Investigating the reproducibility of performing a certain segmentation task across tools and operators, e.g., in a round-robin [38], [39].

Verification of the generated CHP can again include a realism assessment by experts, comparison of organ volumes, weights, and morphological descriptors with corresponding (statistical) information, e.g., from [18], and ascertaining of formal and numerical model correctness (respecting formats, watertight surfaces, etc.; see Section I-B2) “Numerical/implementation considerations” and the study of Gosselin *et al.* [12]. Some of the latter aspects can be automatized and/or integrated as part of the QA process. The verification of the involved steps and tools should be based on identified relevant features that need to be reproduced with sufficient fidelity, as dictated by the context-of-use for the CHP.

b) *Validation:* The assessment of the validation outcome is necessarily application specific and cannot be performed once in a general manner for all possible applications of the CHP. In fact, the suitability of CHPs is also application related. For example, a CHP suitable for traditional ionizing or nonionizing electromagnetic (EM) radiation dosimetry does not necessarily make them suitable for, e.g., biomechanics applications, where information about tissue interfaces (sliding versus connected) and detailed tendon and ligament representations are required, or for ultrasound modeling, where bone heterogeneity representation is crucial. Even for EM dosimetry applications at 5G frequencies, where skin layers become important and need to be correctly represented, typical dosimetric CHPs with simple skin representations might not be suitable. In some cases, the applicability range of CHP can be extended by complementing them with additional information,

such as image-based tissue heterogeneity maps, e.g., the diffusion tensor imaging (DTI)-based anisotropic conductivity maps co-registered to the MIDA model [40], or high resolution subregion models.

Validation by comparison of measurements obtained using an experimental setup representing the use-case measurements with simulation predictions is a compromise between being as close as possible to the real application and context-of-use (to reduce the necessary leap-of-faith and safety margin), and offering optimal experimental control, validation sensitivity, and data richness, at an affordable effort/cost. Validation should be—according to some experts: must be—complemented by an uncertainty assessment, to ascertain the strength of the validation and to judge the agreement between measurement and predictions. For a strong validation, the agreement-success of the chosen experiment must depend on, and hence tests, the model correctness in a highly sensitive manner. The ASME V&V 40 Committee [34] recommends that the rigor of the validation should be commensurate to the risk associated with the context-of-use of the simulation. The risk is a combination of the severity of the consequences from, for example, making an incorrect decision, and the influence the simulation results have on the decision. Another approach to judge the necessary validation rigor is to assess the safety margin between the model-predicted exposure, the danger thresholds, and to demand that the validation be sufficiently robust to keep the confidence interval associated with the model well below that safety margin [41]. Uncertainty assessment should consider relevant factors such as numerical modeling uncertainty, population/inter-person variability, tissue property variability and uncertainty. For many applications involving the human body, the uncertainty about the underlying model equations is also essential, e.g., how well does the commonly applied Pennes bioheat equation (PBE) [42] represent the thermal impact of tissue perfusion and blood flow, even if the parameters were perfectly known. To judge the validation success, the measurement uncertainty typically also needs to be included.

As the purpose of CHPs is to represent the anatomy, validation of the CHP generally necessitates experiments including humans. For example, while experiments on a physical phantom may provide validation evidence for a computational model of dosimetric applications—including simulations that involve a CHP—, they cannot be considered a validation of the CHP for dosimetric modeling. The need to involve human anatomies in CHP validation experiments limits the (ethical) ability to perform such experiments and/or restricts measurement access. An example of experimental validation of CHPs in the context of MRI radiofrequency (RF)-exposure related tissue heating assessment for safety purposes can be found in [43], which also includes a systematic modeling- and validation-uncertainty assessment. In this case, noninvasive means, e.g., surface temperature and MRI-based *in vivo* magnetic field measurements, have been used. Therapeutic applications, such as hyperthermic oncology, which modeling-wise, is similar to the MRI exposure safety case, can justify more comprehensive, invasive validation measurements [44]. Alternatively, cadaver or animal experiments can be used to

validate CHPs, or at least the approach of CHP generation. Recent advances in imaging, e.g., MRI thermometry, increasingly permit extensive and noninvasive data *in vivo* collection, and, hence, richer and more reliable CHP validation.

When creating an entire population of CHPs, validation should also ascertain proper coverage, and statistical representation, of the targeted population. This can involve statistics about the variability of features and comparison with corresponding statistics of the model population, [18], [45], [46], potentially considering correlations.

Validation should not be limited to the CHP, but also encompass associated tools. For example, for a tool to parameterize the BMI or fitness of a CHP by growing or shrinking fat or muscle tissues must be validated.

The issue of validation is also related to the question of CHP detailedness. In various applications, stylized models are in use [47]. However, it typically comes at the cost of increased CHP uncertainty intervals, and hence can result in the necessity for larger safety margins. Once these safety margins have been determined, typically using more detailed and realistic CHPs, they can be factored into safety guidelines and standards, and stylized models can then be used for compliance testing or similar purposes.

*c) Quality assurance:* Complementing the CHP generation process with QA procedures, further increases the trustworthiness of the models. Standardized, and ideally automated, QA procedures can include the following.

- 1) Following standardized segmentation protocols.
- 2) Directives about the tissues to be distinguished and how to deal with ambiguities.
- 3) Examination by an independent expert.
- 4) Assessment of realistic organ/tissue weights and morphologies.
- 5) Checking for unwanted intersections (overlaps) between tissue surfaces, self-intersecting surfaces, holes, or inverted elements (wrong vertex ordering), etc. [12], [35].

*2) Numerical/Implementation Considerations:* In addition to V&V and QA, numerical aspects play an important role in CHP usability and applicability. To increase CHP usability and applicability for simulations, open data formats and compatibility with major software packages are beneficial. For example, Kim *et al.* [28] and Foundation for Research on Information Technologies in Society (IT'IS) [48] made phantoms available in the simple and open surface mesh file format stl (from Stereolithography).

Different considerations apply, depending on whether the region definitions in the model are explicit, e.g., a triangle mesh, or implicit, e.g., based on constructive geometry. To avoid placing a heavy preprocessing burden on users, the following should be considered: all surfaces should be self-intersection free. For models that are to be voxelized, i.e., discretization using a rectilinear, sometimes homogeneous, grid, it is frequently enough to ascertain, “water-tight” surfaces. When unstructured, e.g., tetrahedral, meshing is required, the demands on the model increase considerably: it is very helpful, if the model features compatible (shared) interfaces, if there are no small gaps of thin filling material

layers, and if surface elements are of high quality, i.e., no degenerated, or sliver-like triangles. As the complexity of detailed anatomical models frequently challenges unstructured meshers, it can also be helpful to have multiple versions of the models with varying complexity scales available, such as the V2 models of the ViP [48].

A fundamental aspect relating to modeling the human anatomy using a CHP is to be conscious of common CHP approaches simplifying the anatomy by representing it as homogeneous, neatly delimited regions, when in reality tissue can be heterogeneous, and interfaces can be continuous. To a certain degree, this can be amended by using tissue property maps rather than tissue-specific values during simulations. Furthermore, any CHP involving a segmentation that implies one (or multiple) scale(s) and resolution(s) – typically on an organ/tissue level, while the human body involves a large range of scales, such as the cellular scale.

Part of careful computational modeling is the assessment of solution convergence and mesh independence by performing mesh refinement studies. However, unrealistic tissue/organ surfaces, e.g., stair-cased, or nonsmooth, or too smooth, can converge to a mesh independent solution that is still unrealistic. This occurs particularly when surface effects, e.g., convective surface cooling, or small geometric features are important. The effect is partly related to the model-inherent scale choice and partly a result of the selected model generation procedure. Offering multiple model versions with varying detailedness, as well as employing surface descriptions that permit flexible discretization at varying resolution, can help control the impact of surface discretization.

3) *Tissue Properties*: Predictive simulations of physical (and physiological) interactions with the human body can be achieved only with accurate knowledge of the underlying tissue properties. This section gives an overview of organ and tissue material parameters used in different applications of CHPs, grouped by physics.

The applications driving the development of CHPs include nonionizing EM [49] and acoustic radiation [50], both of which lead to tissue heating [43], as well as ionizing radiation studies [51]. Mechanical tissue properties have been studied extensively, and have been used in conjunction with car crash simulations [52] and other application areas, such as bone fracture risk analysis, and cerebrospinal fluid modeling.

As described in Section I-A, CHPs are 3-D representations (voxels or closed surfaces/solids) of the anatomical structures that separate organs and tissues into individual regions, sometimes referred to as “standard tissues” [53], [54]. These tissue regions are implicitly assumed to be similar in structure and composition, with uniform tissue properties. While this assumption is reasonable in many cases, some tissues clearly have inhomogeneous, or even anisotropic properties, as observed for the anisotropic electrical conductivity of muscle tissues and white matter [55], [56].

Unlike engineered materials, biological tissue is a living material with complex responses that, in many cases, obey laws yet unknown. Living tissues undergo constant change, including relatively fast thermoregulation in response to tissue heating, gradual decrease of water content with aging, and

disease-related responses. There is natural variability among individuals, which in some cases must be accounted for, e.g., in realistic breast imaging studies, where knowledge of the nonuniform breast density and composition are crucial [57].

All tissue properties have a variety of uncertainties associated with the measurement technique, tissue handling, and the environment conditions (e.g., the temperature). In many cases, tissue property measurements in humans, i.e., *in vivo*, are not possible, and values are instead estimated from animal experiments or *ex vivo*. But even when reference measurements for human tissue are available, a certain level of uncertainty remains in regard to the inter-subject variability.

Below we examine the physical tissue properties needed for simulation of physical phenomena in, and around, the human body. Several useful and important resources and databases that provide collections of tissue properties and references are described.

4) *Databases of Tissue Properties*: Tissue properties are usually documented in one of three forms: 1) as a list of measured values; 2) as a functional relationship obtained by fitting the measured values; or 3) as relationships to, e.g., the constituents of the tissue, derived on the basis of an underlying theory. One of the broadest attempts to collect tissue properties for different applications was made by Francis Duck [58],<sup>1</sup> who described tissue properties for mechanics, acoustics, nonionizing and ionizing radiation, and MRI. The CRC Handbook of Mechanical Engineering [59] lists tissue properties and references for several standard tissues, including density, mechanical, acoustic, and thermal properties. For properties related to ionizing radiation, the ICRP has published authoritative references listing tissue composition (constituent elements) and tissue densities for all major organs and tissues [54].

To date, the IT’IS tissue database [60], [61] represents the only continuously updated and curated comprehensive resource. It includes dielectric properties and values of density, heat capacity, thermal conductivity, heat generation rate, transfer (perfusion) rate, low-frequency electrical conductivity (including information about anisotropy in fibrous tissues), viscosity, acoustic propagation, absorption, and magnetic properties ( $T_1/T_2$  relaxation times) for more than 100 biological tissues. Fig. 4 shows an example of thermal conductivity values within the body.

The IT’IS database can be viewed online and downloaded free-of-charge in several human- and machine-readable formats [60] It provides averaged estimates and statistical information about the spread and standard deviation for various thermal, density, perfusion, magnetic properties, acoustic propagation, and low-frequency dielectric parameters. For some tissue properties, e.g., perfusion, the variation can be more than an order of magnitude. The number of data points from which the average for a given property and tissue is calculated varies from 1 to 88 and is provided for each tissue property. Where no values for a standard tissue are available, a proxy tissue is proposed based on consideration of

<sup>1</sup>Unfortunately, Duck’s [58] work has neither been updated since it was published in 1990.




Fig. 4. Thermal conductivity values taken from the IT'IS tissue database [61] depicted in a cross section of a female CHP.

the histological, physiological, and biological composition of the tissue.

While the scientific community has already made an effort to quantify tissue properties accurately, various physical properties exhibit substantial gaps. For example, dielectric tissue properties at low and very high frequencies still have large measurement uncertainties.

#### 5) Tissue Properties in Different Physics:

*a) Mass density:* The mass density  $\rho$ , which is needed for acoustic, mechanical, thermal, EM, and ionizing radiation simulations, is expressed in  $\text{kg/m}^3$ . The density of a tissue can be measured by comparing the mass of a sample in air with the apparent volume displaced in water [58], [62]. A comprehensive and curated collection of tissue mass density values and references can be accessed in the IT'IS tissue properties database [60].

*b) Nonionizing electromagnetic tissue properties:* The dielectric tissue properties from direct current to 100 GHz are best described by the electrical conductivity  $\sigma$  and the relative permittivity  $\epsilon'_r$ .<sup>2</sup> The most comprehensive report of measurements of properties for many tissues was performed by Gabriel *et al.* [56], [63] in 1996. This data was used to fit a frequency-dependent Cole–Cole model, which remains the *de facto* standard for frequencies 1 MHz–1 GHz [64]. Coefficients for the Cole–Cole [65] model and references for evaluating  $\sigma$  and  $\epsilon'_r$  of certain tissues are available as an online resource: 1) IFAC-CNR [66] and 2) the IT'IS tissue database [61].

For frequencies  $<1$  MHz, the measurement technique used by Gabriel *et al.* [56], [63] is known to be less accurate. While other measurements have been performed, low-frequency values are sparse and less reliable with larger uncertainties. Tissue anisotropy at low frequencies further complicates accurate measurement and assignment of tissue conductivity. Realistic anisotropic conductivity distributions can be inferred from DTI and should be included in the CHP, as for, e.g., the MIDA head phantom [40].

At high frequencies, the EM field penetrates only to a narrow region below the body surface. Accordingly, at these frequencies, the skin composition plays a more important role [67], [68].

<sup>2</sup>While  $\sigma$  is the measure of the ability to conduct an electrical current,  $\epsilon'_r$  measure how easily a material polarizes in response to an electric field.

*c) Thermal tissue properties:* Heat transfer in biological tissues is usually modeled with the PBE [42], which describes diffusive heat conduction within a tissue and the magnitude of heat transfer between tissue and blood

$$\rho C_t \frac{\partial T}{\partial t} = k \nabla^2 T + Q + W C_b (T - T_a) + M. \quad (1)$$

The thermal conductivity  $k$  quantifies the ability of a material to conduct heat. The specific heat capacity  $C$  is the amount of heat required to raise the temperature of a tissue by 1 K. Other important properties include the blood-perfusion heat-transfer rate and metabolic heat generation rate  $M$ . Nonmetabolic heat generation sources  $Q$  used in various applications of CHPs include absorbed power from EM radiation to, e.g., investigate tissue heating for mobile phone users and MRI implant safety. Tables with reference values for various thermal tissues properties are available in, e.g., the CRC Handbook of Mechanical Engineering Second Edition [59] or the IT'IS tissue database [60].

While the PBE is clearly most widely applied, it suffers from a range of known limitations, such as insufficient consideration of larger vessels; absence of whole-body thermoregulation and local thermoregulation; absence of tissue anisotropy and inhomogeneity; body-core heating; spatial variability of arterial blood temperature; and tissue damage-related property changes, to name a few. A range of extended or alternative models have been proposed [69], [70] to increase realism or extend the application range. Many of these models, however, require a large number of additional parameters that are rarely available or highly variable.

*d) Acoustic tissue properties:* Acoustic modeling of wave propagation in biological tissues is described by the density and speed of sound  $c$  in the medium and the attenuation of acoustic energy. The absorption coefficient is frequency-dependent and can be modeled with the nonlinear relationship

$$\alpha = \alpha_0 f^b \quad (2)$$

where  $\alpha_0$  is a medium constant and  $b$  is a nonlinearity parameter. Data on nonlinearity is sparse and has been reported for only a subset of standard tissues. Kyriakou [71] suggested that, apart from bone characteristics, acoustic tissue properties of human tissues are poorly quantified. Imaging techniques are more promising and allow acoustic properties to be mapped at high resolution with, e.g., empirical relationships between CT intensity values measured in Hounsfield units and the speed of sound, the density, and the nonlinearity parameter [72], [73].

*e) Other tissue properties:* In contrast to thermal and EM field modeling, ionizing radiation tissue properties are commonly expressed in terms of the elemental composition of the tissues [74]. For a mixture, such as biological tissue, the behavior of photons in the higher EM spectrum (X-ray and gamma radiation) traversing through the medium is described by the mass attenuation coefficient  $\mu$ , which can be approximated closely with the density and mass attenuation coefficients of the constituent elements

$$\frac{\mu}{\rho} = \sum_i w_i \frac{\mu_i}{\rho_i} \quad (3)$$

where  $w_i$  is the proportion of the element by weight. Therefore, the mass attenuation of a specific tissue can be computed as a function of the weight proportion of its constituents.

Computed mass attenuation coefficients for various tissues can be found in [58]. Tables listing the composition and relevant parameters for different standard tissues are published in reports and guidance documents by the ICRP [54].

Tissues exhibit a complex range of mechanical behaviors, including nonlinear, anisotropic, viscoelastic, and, in some cases, also viscoplastic behavior. A detailed discussion of mechanical properties is beyond the scope of this paper, and we refer to other references [59], [75]. Imaging techniques for measuring elastic tissue properties are reviewed in [76].

### C. Libraries of Computational Human Phantoms

Utilizing the techniques described above, many CHPs have been developed and validated for use in the research community. While initially, researchers used simplified or stylized geometries to represent the human body, the growing availability of 3-D medical imaging techniques and high-performance computing infrastructure and software made it possible to create detailed realistic CHPs as seen in Section I-A. In recent years, the development has shifted toward deformation techniques, making it possible to modify existing anatomical models to match certain anthropometric characteristics, such as height, weight, BMI, or organ mass. Deformation techniques allow anatomical shapes to be easily and effectively changed, reducing the time and effort needed to create large cohorts of models [16], [77], [78].

Although the ionizing radiation, nonionizing EM, and automotive communities have been developing similar models with similar tools, there are some important differences and limitations. For example, the ionizing radiation community, guided by the ICRP [54], has placed particular emphasis on creating models with specific organ masses based on the concept of a “reference man” for adults and children of different ages.<sup>3</sup> As a result, most models are not based fully on medical image data of a single subject, but instead have organ masses that are adjusted/scaled to match the ICRP reference model.

The models developed for use by the nonionizing community have been created directly from human image data without modification of organ sizes, e.g., to comply with ICRP. On the contrary, modification of organ sizes without careful modeling of the muscle, fat, and connective tissues surrounding the organs can lead to unrealistic distributions of electrical currents caused by the relatively large differences in electrical conductivities of the various tissue types. Finally, virtual car-crash dummies developed by the automotive industry must allow for realistic mechanical simulations of bone strength and contact response for which, accordingly, the focus has been on accurate representations of cortical bone and ligament function.

The detailed review from 2014 by Xu [6] documents the rapid growth of CHPs (excluding the automotive industry) over the past 50 years. This section provides an update of the developments since 2014, and includes a brief summary of the




Fig. 5. Example libraries of computational phantoms. Models are shown from (a) RPI, (b) UF/NCI, (c) IT'IS, and (d) XCAT series. Only selected phantoms are shown from each population.

phantom libraries currently in use and the key license conditions for many of the well-known phantoms. It also contains some of the model developments that have been underway in parallel in the automotive industry. Due to limited space, several phantom libraries could not be discussed in detail, e.g., [79], [80], and [81]. A more comprehensive list can be found in the review papers by Xu [6], Zaidi and Xu [82], and Zaidi and Tsui [83].

#### 1) Phantom Libraries:

a) *IT'IS virtual family and virtual population*: The ViP library of models, developed in a collaboration between the IT'IS Foundation, the U.S. Food and Drug Administration (FDA), and several other groups, consists of more than ten different anatomical models developed from MRI data of healthy volunteers, including male and female adults and children of various ages, an obese male, an elderly male, a pregnant woman, and newborn [11]. A subset of the ViP models are available in version V2.0 as surface mesh models optimized for FE simulations. The latest version of the ViP, V3.0 released in 2014, is based on higher resolution segmentations ( $0.5 \times 0.5 \times 0.5 \text{ mm}^3$ ) of organs and tissues developed according to strict quality guidelines [12]. To expand the population coverage, and to be able to parameterize certain characteristics (e.g., the body fat content), all phantoms have been made posable and morphable. The morphing and posing approach is based on a realistic biomechanics model to simulate articulation and tissue deformation, expansion, and shrinkage [16].

The ViP models shown in Fig. 5 along with other phantom libraries are widely applied in research and industry. Moreover, they are frequently used in applications to obtain approval from regulatory bodies, e.g., related to implant and MRI safety. As part of more recent developments, in particular in the medical sector, the models are increasingly used to study interactions of electric fields with neurophysiology [84].

b) *XCAT models*: The XCAT models, developed at Johns Hopkins and Duke University, are NURBS-based CHPs able to model cardiac and respiratory motions (RMs) based

<sup>3</sup>Newborn, 1, 5, 10, and 15 years old.



on 4-D tagged MRI and 4-D respiratory gated CT data, respectively [51]. From the initial standard male and female models based on the Visible Human images [17] the XCAT models have been expanded to include a population of anatomically variable phantoms that cover a wide range of height and weight percentiles [18], [19]. The technique is based on an image registration allowing template XCAT anatomy to be mapped to target patient CT data. More recently, volumetric textures that simulate tissue heterogeneity within organs and structures have been added [85]–[87], thereby making it possible to simulate more realistic medical imaging data. Moreover, blood flow models have been embedded to simulate contrast-enhanced imaging [88] and FE models have been investigated to explore variations in cardiac motion [27]. The XCAT phantoms are widely applied to perform virtual clinical trials (VCTs) in medical imaging research, to simulate radiation doses from radiography and radiotherapy, and to improve imaging devices and techniques [89].

*c) GSF voxel phantoms:* The models generated at the German Research Center for Environmental Health (GSF, now the Helmholtz Institute) consist of a collection of 12 voxel phantoms based on CT data of living patients. One exception is the eight-week-old baby model generated from image data of a dead infant [90]. The group includes adult male, female, pediatric, and pregnant-woman phantoms.

The GSF voxel phantoms were originally developed to be used for radiation transport calculations in clinical dosimetry, environmental, and occupational exposures, from external and internal sources. Therefore, the GSF voxel phantoms contain a large number of organs and tissues, including most of the ICRP critical organs.

*d) RPI library of models:* Various models were developed at the RPI, including VIP-Man [91], which was the first CHP created from the Visible Human image data [92] and the RPI male and RPI female models, which have both been carefully adjusted to match ICRP Pub. 89 reference organ masses [77]. These two reference models were later scaled to create different percentiles of weight-specific phantoms. The RPI models are commonly used in studies related to ionizing radiation. The RPI collection of models also include a pregnant female model at different gestational phases.

*e) University of Florida Family of Models:* The University of Florida (UF) “family of models,” developed at the UF and the National Cancer Institute of the National Institutes of Health (NIH/NCI), is a series of hybrid CHPs that includes models of an adult male, adult female, pregnant female, and pediatric models of various ages (newborn to 15-year-old) and both genders [93]–[95]. Assembled from CT image data obtained from both cadavers and living patients, the library covers variations in body size and shape among the U.S. population of adults and children.

The UF hybrid phantom series are available as NURBS and polygon mesh surfaces (hybrid-NURBS/PM phantoms) that allow users to apply morphing techniques to develop nonreference phantoms.

*f) HDRK-Man and HDRK-Woman:* The high-definition reference Korean-man (HDRK-Man) and Korean-woman

(HDRK-Woman) are models that have been developed by researchers at Hanyang University in Korea [96], [97]. These models have been generated from the segmented Visible Korean Human cryosection image data [8]. The quality and resolution of the data ( $0.2 \times 0.2 \times 1.0 \text{ mm}^3$ ) allow accurate segmentation of organs and tissues. These models, originally developed for radiation protection purposes, have been adjusted to match the reference Korean data, including height, weight, and organ masses.

*g) Open source models:* AustinMan and AustinWoman, based on the Visible Human data [92], are open-source voxel models developed at the University of Texas at Austin [98] with a focus on EM simulations. The models are available under the Creative Commons license with a noncommercial, no redistribution clause [99]. A female and child models, which are being actively developed by the European projects PIPER [100] and ViVa [101] for car-crash safety applications, are available under open source licenses (with a noncommercial clause).

*h) Models developed for the automotive industry:* There are several efforts in the automotive industry to develop CHPs for car-crash safety simulations. The Global Human Body Models Consortium is an international consortium of seven different partners, most of whom are automobile manufacturers [102]. The consortium is working on the development of both male and female CHPs of different heights<sup>4</sup> in both standing and sitting positions [103]–[105]. The Total Human Model for Safety is a family of models developed by Toyota that include a fifth percentile adult female, 50th and 95th percentile adult males, and 3-, 6-, and 10-year-old children [106], [107].

*2) Availability of CHPs:* Many of the well-known CHPs are owned by their creators, usually a research institute [12], [18], [95]. In some cases, a consortium has joined forces to share the effort and costs of developing new models [102]. There are also several open source projects [78], [98], [100], which try to establish wider community access and involvement. While access to the models is typically granted to collaborators, license conditions for academic and commercial use are not always clear. Table I provides an overview of the license conditions for commonly used CHPs. Most CHPs can be obtained free of charge for academic use. For commercial use, many CHPs are available for a fee, while for some CHPs commercial applications are prohibited. The original image data used to create the phantoms is in most cases not available, with the exception of the phantoms derived from the widely used Visible Human [92] and Visible Korean data [8].

## II. APPLICATIONS OF REALISTIC COMPUTATIONAL HUMAN PHANTOMS IN BIOMEDICAL ENGINEERING

The libraries of CHPs discussed above, and countless others, have been utilized for many applications, including ionizing radiation dosimetry and imaging, nonionizing EM radiation leading to tissue heating and neurostimulation, and passive

<sup>4</sup>5th, 50th, and 95th percentiles.

TABLE I  
SOURCE AND AVAILABILITY OF CHPs

Name, owner	Format	Academic use	Commercial use	Image data	Reference
ViP V1.x, IT <sup>2</sup> S Foundation, Switzerland	voxels proprietary (Sim4Life <sup>1</sup> )	free	license fees may apply	MRI	[11]
ViP V3.x, IT <sup>2</sup> S Foundation, Switzerland	proprietary (Sim4Life <sup>1</sup> )	free	license fees may apply	MRI	[12]
XCAT, Duke University, USA	NURBS surface mesh voxels	license fees may apply	license fees may apply	NLM Visible Human chest CT used for personalization modified to ICRP-89 reference	[18]
GSF Voxel Phantoms, Helmholtz Institute, Germany	voxels	free	prohibited	CT modified to ICRP-89 reference	[90]
RPI Phantoms, Rensselaer Polytechnic Institute, USA	surface mesh voxels	free (collaborators)	n/a	NLM Visible Human CT modified to ICRP-89 reference	[77]
UF Library, University Florida/ NIH NCI, USA	NURBS surface mesh voxels	free	license fees may apply	CT modified to ICRP-89 reference	[93]
HDRK, Hanyang University, South-Korea	voxels surface mesh tetrahedral mesh	free	unknown	Visible Korean modified to reference Korean	[96]
Japanese Voxel Phantoms, Japan Atomic Energy Agency	voxels	unknown	unknown	CT modified to ICRP-89 reference	[80] [211]
Chinese Voxel Phantoms Institute for Radiation Protection, China	voxels	unknown	unknown	Visible Chinese	[79] [210]
Pediatric Phantom Library Geneva University Hospital	voxels	unknown	unknown	modified from Virtual Population pediatric models	[81]
GHBMC, GHBMC/ Elemance, USA	voxels surface mesh proprietary (LS-DYNA, PAM-CRASH, ...)	license fees may apply	license fees may apply	CT and population statistics (e.g., cortical thickness)	[102] [212]
THUMS, Toyota, LSTC, USA	proprietary (LS-DYNA)	unknown	unknown	CT	[106]
AustinMan/Woman, University of Texas at Austin, USA	voxels	free CC BY-NC-ND 3.0 [99]	prohibited CC BY-NC-ND 3.0 [99]	NLM Visible Human	[98]
PIPER project, Open Source	proprietary (LS-DYNA)	free GPLv3	free GPLv3	unknown	[100]

<sup>1</sup> Sim4Life is a commercial physics simulation platform for computational life-science, [www.zurichmedtech.com](http://www.zurichmedtech.com).

car safety. The following section highlights the many uses of CHPs within various fields of research.

#### A. Applications to External and Internal Dosimetry for Ionizing Radiation

The assessment of radiation absorbed dose to the internal organs and tissues of patients in the field of biomedicine has relied heavily upon the development and application of CHPs. With only few exceptions (OSLD, MOSFET, and radiochromic film measurements on the patient's body surface, for example), direct measurement of radiation dose to exposed tissues during medical imaging or radiotherapy is not feasible, and thus tissue dosimetry is assessed computationally through various dose algorithms. Increasingly, however, MC transport techniques are used with CHPs of the patient to assess organ dose. These CHPs can range in both format type—stylized, voxel, or hybrid—and morphometric category—reference (small library of average height/weight individuals), patient dependent (expanded library of variable height/weight models), patient sculpted (model morphed

to match the patient's outer body contour), or patient specific (model developed from the patient's own CT or MRI images) [108]. The ultimate goal of patient dosimetry in medical imaging would be the use of a patient-specific CHP. However, this is not clinically feasible given current limitations in organ segmentation algorithms, and one must instead select, and then potentially sculpt, an existing CHP that best morphometrically matches the given patient.

Another use of CHPs for medical dosimetry is to explore changes in radiation dose with variations in imaging or therapy technique factors (x-ray entrance fluence, x-ray beam quality, or level of injected activity, as examples) across different populations of patient morphometry (height, weight, and BMI) and possibly organ morphometry (organ size, shape, and depth in the body). In the vast majority of cases, it would be impracticable and potentially unethical to perform these dose sensitivity studies in a clinical trial using real patients. However, with the scalability and morphing capabilities of CHPs, digitally based studies of dose sensitivity are feasible.

The following section summarizes selected articles covering a broad range of medical dosimetry applications which




Fig. 6. Boxplots comparing all organ dose percent differences for each of the six matching parameters. The vertical lines extend at most 1.5 times the interquartile range.

highlight the utility and applications of difference CHP libraries.

In 2012, Ding *et al.* [109] reported on the development of an extended CHP library based upon the RPI-adult male and female, and their use in quantifying CT imaging dose with changes in body size. Ten phantoms were developed with BMIs ranging from  $23.5 \text{ kg/m}^2$  (normal weight) to  $46.4 \text{ kg/m}^2$  (morbidly obese). CT imaging doses were assessed for their computational source model of a GE LightSpeed 16-slice scanner operated at 120 and 140 kVp. With a constant mAs, radiation doses to abdominal organs such as the colon were found to be up to  $\sim 60\%$  smaller for the obese compared to the normal weight phantoms. Increasing the tube potential from 120 to 140 kVp for the same obese individual resulted in increased organ doses by as much as  $56\%$  for organs within the scan field, such as the stomach, and  $62\%$  for those out of the scan field, such as the thyroid. These phantoms have been since integrated within the commercial software package VirtualDose [110].

In 2017, Stepusin *et al.* [111] assessed six different patient-to-phantom matching criteria for CT dosimetry. These criteria included: 1) age-matching to reference phantoms; 2) height/weight matching to members of the UF/NCI hybrid phantom library [112]; 3) effective diameter matching as assessed in the central slice of the scan range; 4) effective diameter matching as averaged across the entire scan range; 5) water equivalent diameter matching as assessed in the central slice of the scan range; and 6) water equivalent diameter matching as averaged across the entire scan range (see Fig. 6). The study was conducted using two phantom libraries—a series of 52 patient-specific voxel phantoms (representing “real” patients) and the 351-member UF/NCI library of hybrid phantoms. The study concluded that while water equivalent diameter matching is superior in terms of dose accuracy, height-and-weight matching is an acceptable and reliable method for matching patient to a member of a CHP library for CT dosimetry.




Fig. 7. DAHs for ten select high-dose cases normalized to peak skin dose. Ordinate indicates what fraction of peak skin dose is delivered to an area of exposed patient skin given on the abscissa.

In 2017, Fu *et al.* [113] reported a study in which 21 members of the XCAT CHP library [18] were used to compare organ doses delivered by the GE Revolution CT scanner under either tube current modulation or organ-dose modulation (ODM). In the XCAT phantom series, two breast glandularities (ratio of glandular to adipose tissue mass) were modeled at 50/50 and 20/80. ODM was found to significantly reduce organ doses.

In 2017, Borrego *et al.* [114] used the UF/NCI CHP library to explore patient skin and organ dose in fluoroscopically guided interventions. Ten high cumulative reference air kerma radiation dose structured reports were selected for dosimetric analysis as applied to height-weight matched members of the CHP library. The RIPSA algorithm was applied, demonstrating the ability to report dose-area histograms (DAHs) of skin dose for each individual patient [115] (see Fig. 7). Detailed dosimetric analyses of two patients undergoing a bilateral uterine artery embolization and an abdominal angiography were performed, demonstrating that individual irradiative events<sup>5</sup> provide contributions to cumulative organ dose that vary by up to three orders of magnitude. These results provide for the possibility of truncating the number of MC simulations required with minimal impact on dose accuracy.

In 2017, Zvereva *et al.* [116] reported on a feasibility study to reduce differences in diagnostic nuclear medicine doses that are traditionally based upon MC simulations of photon and electron transport in an age-matched reference phantom. The study focused on the positron emission tomography (PET)-based tumor imaging agent (S)-4-(3- $^{18}\text{F}$ -fluoropropyl)-L-glutamic acid, and used the MIRD schema for dose assessment [90]. The study employed two hypothetical patients selected from the HMGU (Helmholtz Zentrum Munchen voxel phantom library): the 103.2-kg male Visible Human phantom (patient #1) and the 51.5-kg Irene female phantom (patient #2) [90] (see Fig. 8). Organ doses were

<sup>5</sup>117 and 299 events, respectively.




Fig. 8. Polygon mesh version of the ICRP 110 reference phantom with selected dimensions and the result of its adjustment to patients #1 and #2.

further assessed in the ICRP publication 110 reference computational phantoms in both their voxelized [30] and polygon-mesh formats [117]. Finally, the polygon-mesh phantoms were sculpted to match either patient #1 or #2 through a multistep process to include separate and proportional scaling of the legs, torso, and arms. Importantly, no further adjustments were made to individual organ size, shape, and depth to “match” the targeted hypothetical patients. The study concluded that it is feasible to estimate patient-specific nuclear medicine organ doses within a relative uncertainty of 25% or less using phantom sculpting to the outer body dimensions and shape of the nuclear medicine patient.

In 2017, Petroccia *et al.* [118] explored the use of the UF/NCI computational phantom library [112] to represent patients historically treated with cobalt-60 gamma-rays for the treatment of Hodgkin’s lymphoma. This patient population is ideal for radiation epidemiology studies of late clinical effects for both secondary cancer and cardiac disease. To perform dose-response studies, however, organ doses—both in and outside the treatment field—must be obtained. In this paper, the authors looked at patients treated in the early 1960s to early 1970s in which only patient gender, age, height, and weight were available, along with treatment planning information. In this paper, the authors developed a detailed workflow resulting in a complete radiation treatment model of the historical radiotherapy using height/weight-matched CHPs from the UF/NCI series. Organ doses could be reported for the first time for these patients by individual treatment field, as well as cumulative over the full course of radiotherapy.

Papadimitroulas *et al.* [119] published in 2018 a new methodology to estimate the absorbed dose in organs. The method is based on specified clinically derived radiopharmaceutical biodistributions and personalized anatomical characteristics. The authors use realistic MC simulations and pediatric CHPs to calculate a parameter called the “specific absorbed dose rate” (SADR), i.e., a unique quantitative metric which is specific to a particular organ. The SADR is defined as the absorbed dose rate in an organ when the biodistribution of radioactivity over the whole body is considered. SADR enhances the personalization of dosimetry in pediatric nuclear medicine and provides time-dependent organ dose rate curves.

As a final example, we look at a 2012 study by Senthamizhchelvan *et al.* [120] who explored tumor dosimetry

and response for  $^{153}\text{Sm}$ -EDTMP (Quadramet) therapy in patients with high-risk osteosarcoma. In this paper, radiation absorbed dose in 19 tumors within six osteosarcoma patients were treated with  $^{153}\text{Sm}$ -EDTM. The dose was assessed within the 3-D-RD software package in which the CT portion of the SPECT/CT patient image is used as the patient-specific voxel phantom and the SPECT portion is used to define, at the voxel level, the time-integrated activities of the radiopharmaceutical. Patients were treated first, with a low-dose protocol (37–52 MBq/kg), then following bone marrow recover, by a high-dose protocol (222 MBq/kg). Tumor doses determined by MC radiation transport within 3-D-RD were determined and compared to doses reported by the spherical tissue models of the OLINDA nuclear medicine code [121]. Mean tumor doses were reported to range from 1.5 to 15 Gy for the low-dose therapy, and from 3 to 60 Gy for the high-dose therapy. Mean tumor absorbed dose reported by the OLINDA code were within 5% of the mean values reported by 3-D-RD.

### B. Applications to External and Internal Dosimetry of Nonionizing Radiation

Research on human exposure to EM fields and the development of safety guidelines and compliance standards began in the second half of the 20th century [122]. At the beginning, due to lack of computational resources, only very generic representations of the body, such as prolate spheroids [123] or 1-D stratified representations of the tissue structures [124], could be used. Nevertheless, these simple models provided fundamental insight for the assessment of the whole-body absorption and local absorption in the far-field [123], and in the near-field [125].

In the development of safety guidelines [126], [127], defined in terms of basic restrictions (induced fields) and reference levels (incident fields), CHPs are applied with the objectives to: 1) identify the key anatomical parameters for quantification of the fields absorbed in the body and 2) correlate the external fields and the absorbed fields in terms of an exposure mechanism.

For compliance testing [128]–[130], anatomical models are applied for: 1) the development and validation of CHPs for testing against the basic restrictions and 2) statistical evaluation of the basic restrictions by means of numerical simulations of a representative set of CHPs.

In cell phone compliance testing, the correct rendering of the proportions of the human head is a key parameter, because the exposure of tissue in the close environment of an EM field source strongly depends on the distance of the exposed tissue from the source [125], [131]–[133]. Simulation results of detailed CHPs of adults and children [131], [134], [135] served as basis for the design of the dosimetric head phantom that is now used in cell phone compliance testing standards [128], [129]. It has been continually validated versus the latest available anatomical head models of adults and children [136]–[138] (see Fig. 9).

A similar approach was chosen for the compliance testing of base station antennas [139], where a large number of whole body CHPs [11] in different exposure situations were used to




Fig. 9. Tissue losses, i.e., absorption in a cross section of DUKE (member of ViP), and the E-field lines generated by the transmitter in the vicinity of the phone.

develop a heavily simplified body phantom that takes height, BMI, and tissue layering of the exposed population group into consideration [140].

For low-frequency applications, such as resonant wireless charging or induction cooking [141], [142], there are no available standardized experimental phantoms. As the reference levels may be exceeded in the close near field of these applications [126], [127], compliance with basic restrictions is tested with CHPs.

Although there are practically no alternatives to the numerical assessment of the fields induced in the body in the low frequency range, standardization efforts for numerical and combined numerical and experimental methods have so far focused on the RF range. The IEC/IEEE 62704 series of standards gives guidance on the numerical assessment of the peak-spatial specific absorption rate based on both body phantoms and anatomical models [143], [144].

The safety standards for MRI scanners IEC 60601-2-33 [145] are currently also revised based on advanced modeling techniques combined with thermal modeling using perfusion tissue models [146], [147]. ISO/TS 10974 standard [148], defines a method based on CHP to demonstrate MRI safety of active implants [149] (see Fig. 10). The shortcomings of the experimental approach described in [150] have been largely overcome with ISO/TS 10974 [148], which allows for significantly more realistic modeling of the tissue heating due to induced currents on the implant and for the statistical risk evaluation for particular patient groups and implant types.

With the latest wireless devices technologies, frequencies above 10 GHz are applied, whereby the penetration depth of EM fields in the body drops to a few millimeters or less. In the upcoming revisions of the ICNIRP guidelines [126] and IEEE




Fig. 10. Induced absorption of the B1-field in DUKE (ViP) during MRI scans (right side) and the associated local tissue temperatures (left side).

C95.1 [127], basic restrictions will be introduced in terms of the transmitted power density. The low penetration depth poses additional demands on anatomical modeling as the skin can no longer be regarded as bulk tissue, i.e., the distinct dielectric properties of its layers need be considered [151], [152].

### C. Radiation Protection

The system of radiation protection is built upon several principles: its main aim is to limit the potential health detriment due to exposure to radiation below certain acceptable limits. The central assumption of a linear dose-response relationship for the induction of cancer and heritable effects directly connects the limitation of the potential health detriment with a limitation of the dose to the human body from radiation. Since radiation doses in the human body cannot be measured directly, a complex system involving various types of quantities has been established. So-called “operational” quantities, such as the personal dose equivalent  $H_p(d)$  and the ambient dose equivalent,  $H^*(d)$  [153] can be measured by personal or environmental dosimeters. In contrast to the “protection quantities” in the human body: the organ equivalent dose  $H_T$  and the effective dose (ED) [154], where the latter is a weighted average of organ equivalent doses reflecting the different radiation sensitivity of individual organs and tissues of the body. Since the system of radiation protection is also of a prospective nature, the dose quantities are not determined on an individual basis, but consider rather whole populations.

Due to these principles, CHPs play a central role in radiation protection dosimetry. The first CHPs used for this purpose were based upon mathematical expressions representing planes, cylindrical, conical, elliptical, and spherical surfaces describing the shape and position of idealized body organs [4], [47], [155], [156]. For this first generation of CHPs, the organ masses and volumes were in accordance with the ICRP data of former reference man [53].

With the advent of more powerful computers in the 1980s, various groups have developed voxel phantoms based on 3-D images as an extension and improvement to these earlier models [82], [90], [91], [157]–[167]. Various authors have shown that the organ shapes of the earlier mathematical phantoms present an over-simplification that has an influence on the energy distribution, which may deviate for some cases systematically from that calculated for voxel phantoms. For external radiation, the parameters influencing the organ doses are: 1) the depth of the organ below the body surface; 2) the exterior shape of the trunk; and 3) the trunk diameter relative to the incoming radiation beam [168]–[172]. For internal dosimetry, the influencing parameters are: 1) the relative position of source; 2) the target organs (for organ cross-fire); and 3) the organ mass (for organ self-absorption) [173]–[176].

However, most of these CHPs do not represent the average Caucasian man or woman, due to being derived from a specific individual. Hence, the ICRP decided to construct voxel CHPs being representative of the adult reference male and reference female [54] with respect to their external dimensions, their organ topology, and their organ masses [30], with the organ dose conversion coefficients following the recent ICRP recommendations [154].

Two male and female voxel CHPs were selected with external dimensions close to the reference data [167], [177]. These two CHPs were then adjusted to the reference values of ICRP Publication 89 [54] in several steps.

- 1) Adjustment of the body height and the skeleton mass by voxel scaling.
- 2) Adjustment of the individual organ and tissue masses to the reference values by adding or subtracting the required number of voxels.
- 3) Additional organ and body region modifications.

The bones were subdivided into a cortical shell and a spongy region accommodating the bone trabeculae together with the marrow cavities—structures that are both much smaller than the voxel resolution of the phantoms. The method of constructing the adult male and female reference computational phantoms is described in detail in ICRP Publication 110 [30].

The adult reference CHPs of ICRP Publication 110 are the official CHPs representing the reference male and reference female [54], [154] (see Fig. 11). Each of these reference CHPs has 140 different organs and tissues. The male phantom consists of approximately 1.9 million voxels with a resolution of  $2.137 \times 2.137 \times 8.0 \text{ mm}^3$ , and the female phantom consists of approximately 3.9 million voxels at a slightly finer resolution of  $1.775 \times 1.775 \times 4.84 \text{ mm}^3$ . The ICRP has recently published (Pub. 116 [178] and Pub. 133 [179]) recommended values for dose coefficients for both external and internal exposures using these two phantoms.

Although these phantoms have reference organ masses, they still have individual organ topology reflecting the tomographic data used for their construction. Obviously, both models cannot represent real individuals, and thus they should not be used to assess doses for specific individuals.

Further limitations of the reference CHPs are due to their voxel resolution of the order of millimeters, specifically concerning the slice thickness of several millimeters, which results



Fig. 11. Adult male and female mesh-type ICRP reference computational phantoms [180].

in stepped organ surfaces and nonclosed contours for skin and hollow organs, coarse and unrealistic representation of the eye lenses, and the inability to represent very fine structures, such as some specific source and target regions in the alimentary and respiratory tracts having micrometer dimensions. This resulted in the necessity to use several stylized CHPs in addition to the voxel CHPs for the calculation of reference dose coefficients for high-resolution source and target regions and radiations of low penetration.

These limitations are being addressed by a research project for creating BREP versions of the reference CHPs. The resulting mesh-type reference CHPs closely resemble the ICRP Publication 110 [30] phantoms in their main anatomical features and thus meet the requirement of being “exact replicas.” However, they include micrometer-fine source and target regions that could not be represented in the Publication 110 phantoms. The methods applied to create these mesh-type reference CHPs [180] are similar to those described in Section I-A1.

#### D. Biomedical Imaging

CHPs serve as virtual patients within biomedical imaging simulation studies. They are imaged by means of accurate simulation methods, e.g., MC or analytical, with models that emulate the characteristics of the particular imaging modality, instrumentation, and system configuration, as well as the physics of the imaging process including all artifacts. The simulation codes generate imaging data that can be reconstructed and analyzed on clinical workstations [86]. Through the combination of CHPs with medical imaging simulation methods, realistic biomedical images that closely mimic those acquired




Fig. 12. CT simulation using an CHP. Imaging data is acquired from the CHP using the scanner model; the data is then reconstructed into the simulated CT images.

from live subjects can be generated entirely on the computer, as shown in Fig. 12.

Simulated images from the CHP include the image degradation effects by the imaging system and the imaging process. They can be compared to the original phantoms, which represent the known “truth,” and the difference can provide quantitative evaluation of the imaging instrumentation, the image processing, and reconstruction methods. The ultimate goal is to improve image quality for more accurate detection of abnormalities and diagnosis of diseases.

With this ability, CHPs are able to assess the effects of anatomical, physiological, physical, and instrumental factors on the imaging process, and to study image acquisition strategies, processing and reconstruction methods, and visualization techniques. In this section, we present examples to illustrate the applicability of CHPs for biomedical imaging.

1) *Effect of Noise or Artifacts of the Imaging Process:* Within medical imaging simulation, one can easily alter system design parameters and acquisition techniques and witness these effects on the resulting images as compared to the known phantom. Different effects can be singled out and studied independently of other potential sources of artifacts. Fig. 13 shows examples of cross-sectional CT images using the XCAT series of CHPs showing the effects of noise and beam-hardening.

2) *Optimization and Comparison of Modalities:* CHPs can be imaged repeatedly to find the system values or techniques that give the highest image quality. In addition, the radiation dose to the organs and structures can be calculated to assess the radiation exposure from different imaging systems and techniques. CHP's therefore provide the basis to optimize and compare imaging applications in terms of image quality and dose.

Fig. 14 shows a comparison of chest CT, tomosynthesis, and radiography in terms of the ED in patients of various sizes. Fifty-nine anatomically variable CHPs were imaged with each




Fig. 13. Top: sample transaxial-slice CT images of the head at 80 kVp (left) and 120 kVp (right) showing a beam-hardening artifact (arrow). Bottom: sample noisy transaxial-slice abdominal CT images at 1 mAs/view (left) and 0.1 mAs/view (right).




Fig. 14. ED plotted as a function of patient size for CT, tomosynthesis (Tomo), posteroanterior + left lateral radiography (PA + Left LAT), and anteroposterior radiography (AP).

modality and the organ doses and EDs were estimated using MC simulations [181]. It can be seen in the figure that CT, as expected, has the highest dose, followed by tomosynthesis and radiography. Patient body size can also be seen to have more of an impact on the CT dose, meaning that patients of larger size may benefit from choosing tomosynthesis over CT.

Studies such as this, which involve imaging patients with different modalities or imaging multiple times with varying parameters, would be ethically and practically impossible to perform using live subjects.

3) *Development of Image Processing and Reconstruction Methods:* Realistic CHPs have played an important role in the development of image reconstruction methods compensating for artifacts. Fig. 15 shows an example of the evaluation of a SPECT image reconstruction method that provides significant improvement in both image quality and quantitative accuracy, as seen in a comparison to the “true” radioactivity distribution modeled in the particular phantom.




Fig. 15. Simulated myocardial SPECT images reconstructed using an iterative order-subset expectation-maximization (OS-EM) method with accurate models, i.e., the XCAT phantom, of the system response and imaging physics for significant improvement in both image quality and quantitative accuracy. From top to bottom row, OS-EM images obtained without any modeling (first row), with the collimator-detector response (CDR) (second row), with CDR and attenuation (third row), and with CDR, attenuation, and scatter modeling (fourth row). From the left to right column: OS-EM with 1, 2, 3, 4, 5, 10, and 20 iterations.




Fig. 16. Sample noise-free transaxial-slice cardiac CT images at mid-diastolic phase with fast full scan (left) and at 333 ms/rotation short scan (right) showing the effect of motion of the coronary artery (arrows). Images were sharpened to show the artifacts more clearly.

4) *Effect of Cardiac and Respiratory Motions:* Motions such as the cardiac, respiratory, and patient voluntary motions can cause artifacts in medical images which can lead to misdiagnosis. CHPs can be used to simulate different types, and extents, of motions artifacts, e.g., Fig. 16 shows the effect of cardiac motion on CT imaging of the coronary arteries. Streak artifacts can be seen on the arteries that move the most while the heart beats.

Fig. 17 shows the use of simulation methodologies demonstrating cardiac SPECT imaging artifacts due to respiratory heart motion. The heart moves up and down with the diaphragm, causing artifacts mimicking defects, or areas of cardiac tissue receiving less blood perfusion.

5) *Motion Compensation:* Methods for motion artifacts include various acquisition strategies such as gating, patient instructed breathing, and 4-D reconstruction algorithms. CHPs and simulation techniques provide a nice toolset to quantitatively evaluate and optimize these methods.

In Fig. 18, we demonstrate the application of the 4-D XCAT phantom in the development and evaluation of a 4-D RM




Fig. 17. Simulated myocardial SPECT images demonstrating the effect of RM (left column). Same sample short-axis slice through the center of the heart (right column). From the 4-D XCAT phantom (top row) without RM, and with (middle row) 2-cm and (bottom row) 4-cm RM amplitude showing the increasing motion artifacts on the superior and inferior wall of the myocardium.




Fig. 18. Evaluation of a 4-D PET image reconstruction method with RM compensation using the 4-D XCAT phantom with RM capabilities. (a) Sample coronal slice from the 3-D XCAT phantom with three lung lesions and no RM. The PET images of the same coronal slice obtained from (b) 3-D maximum likelihood-expectation maximization (ML-EM) image reconstruction algorithm with no RM compensation, (c) 4-D ML-EM image reconstruction with RM compensation using known RM modeling, and (d) 4-D ML-EM image reconstruction with RM compensation using an estimated 4-D RM model.

estimation technique and a 4-D PET image reconstruction method with accurate motion compensation. The simulated images show the effect of RM, blurring, and thereby decreasing, the lung lesion detectability. It also demonstrates the ability of the 4-D image reconstruction method, with accurate motion modeling and compensation, to provide improved lesion detection.

### E. Regulatory Applications

CHPs currently play an increasingly significant role in the regulatory approval process for medical devices by helping characterize the safety of MRI systems and devices entering the MRI environment. CHPs have the potential to significantly impact future regulatory approvals, particularly in the context of VCTs, in which *simulated* human clinical trials (HCTs) are performed *in-silico*. VCTs are part of a broader *in-silico* clinical trial framework [182] where computer simulation data, i.e., digital evidence, are used for regulatory decision-making along with traditional sources such as bench testing, animal testing, HCTs, and adverse event reporting. Currently HCTs serve as the gold standard, providing reasonable assurance of safety and effectiveness for high-risk medical devices. This is mainly because HCTs allow for evaluation of complex biological, chemical, and physical interactions between the human body



and a medical device. As computational modeling techniques improve, and trust in the predictive capability of computational models grow, the reliance on HCTs may be reduced as they are augmented by digital evidence from *in-silico* VCTs.

During the life cycle of a medical device, from ideation, to development, regulatory approval, and ultimately real-world clinical use, evidence regarding device performance in the intended patient population is essential for ensuring public health. Each of the sources of data, i.e., bench, animal, HCT, etc., has advantages and limitations [183]. Therefore, the appropriate combination of data from each source, particularly in the context of regulatory approval depends on the risks in the intended use for the device. Evaluation of a low-risk device may primarily rely upon bench testing and adverse event reporting, while for a high-risk device, evaluation may require HCT data. A major advantage of HCTs is the assessment in actual patients, which provides assurance that the data will be indicative of clinical use. However, significant barriers broadly impede the use of HCTs in device evaluation. Cost and patient safety are frequent concerns; recruitment of patients is required, and significant demands may be made upon physicians' time and a hospitals' resources (e.g., Institutional Review Boards). HCTs are usually costly and lengthy, potentially taking years to complete. They may also expose patients to novel or unproven technologies that pose a risk of harm.

*In-silico* clinical trials have become widely adopted for regulatory use in the evaluation of RF-induced heating during MRI. The RF fields created by MRI systems deposit energy into the patient, generating heat, particularly if electrically conductive implants are present, that can potentially cause tissue damage [184]. This is an important safety concern that should be addressed, both for MRI systems [185] and MRI compatible medical devices [186]. Implantable medical devices are concerning because they could potentially act as antennas, focusing energy deposition, resulting in thermal tissue damage. The amount of energy deposited and how it is distributed within the patient is related to the patient's anatomy, weight, MRI landmark positioning, and the location and orientation of the implant(s). Testing in simplified physical phantoms cannot adequately capture this variability, whereas evaluation of potential RF heating in patients is not ethical, or even feasible. Thus, there is a strong regulatory need for *in-silico* clinical trials that can provide alternative data. That need is currently being met by CHPs such as the virtual family [11] and the VIP [12]. Finite difference time domain simulations allow for localized estimates of energy deposition and temperature rise [187], which can be used to determine the worst-case scenario in the patient population and help ensure MRI systems, and implants, will not exceed safe temperature limits in real patients.

A potentially important future role of VCTs is performance evaluation of imaging systems. Such a virtual evaluation typically involves CHPs, a model of the device physics, the image reconstruction process, and a model for image interpretation for a specified clinical task. As outlined below, this is an active research area, particularly for breast imaging devices such as mammography and digital breast tomosynthesis (DBT),

a limited-angle tomographic imaging system. These devices use ionizing radiation to screen millions of women per year, thus safety and efficacy are significant public health concerns. Currently DBT devices are considered high risk devices (class 3) in the U.S., indicating the need for HCTs. Conducting an HCT for DBT devices can however be burdensome due to the use of ionizing radiation and periodic design changes. Bench testing with physical phantoms is routinely performed, however the complex and variable texture of breast images can make it difficult to infer clinical performance from simple phantom-based measurements alone, particularly when there are nonlinearities in the imaging system. The desire to limit patient radiation exposure and noted inadequacies of bench testing has motivated the development of VCTs and associated breast CHPs, targeted at both technology development and regulatory approval.

Virtual imaging clinical trials for cancer screening devices typically assess the detectability of cancerous lesions and require modeling of the imaging process from the patient to image interpretation. A CHP population consisting of cancer positive and negative patients is required. The population must be imaged virtually using physics-based simulation of the image acquisition process for each device in the study. Images are then reconstructed and interpreted by a human, or algorithm, to determine some measure of cancer detectability, thus simulating a breast cancer screening HCT. Anthropomorphic breast CHPs are a vital component of these VCTs due to the influence of surrounding anatomical structures on lesion detectability. Breast phantoms generally follow one of two design philosophies, segmentation of 3-D breast CT data [188], [189] or procedural generation of random breast tissue structures [190]–[193]. Segmentation of CT images allows for accurate representation of large scale breast anatomy, but CT image resolution does not currently allow for segmentation of fine structures such as ligaments that are visible in higher resolution mammographic and DBT systems. The size of the VIP is also limited to the number of CT data sets available, though some efforts have been made to solve this problem with eigenfunction analysis [194]. Procedural generation on the other hand allows for a virtual patient population of arbitrary size at an arbitrary voxel resolution, but current phantoms generated in this manner have limitations in terms of realism. A combination of segmentation and procedurally generated fine structures has been proposed to provide some of the advantages of each technique [195]. Models for cancerous breast lesions have also been developed [196], [197]. Several research groups have made significant progress in creating breast VCT platforms and have used them for performance evaluation studies [198], [199]. An open source software framework for managing VCT data has also been made available [200], [201]. VCTs are not currently used for regulatory evaluation of imaging systems, but the FDA has commenced the research project Virtual Imaging Clinical Trials for Regulatory Evaluation [202] to study the issues with the hope to adopt VCT for regulatory use in the future.

In a VCT, virtual patients are enrolled, and analogous to an HCT, where the interaction between the device and the virtual patient is studied. This process typically involves multiple

inputs (CHPs and device parameters) and computational models. Increased use of VCT data for regulatory decision making will certainly be contingent upon proper V&V of the CHP and the different VCT models. As previously described, V&V of a computational model is “the process of determining the degree to which a model is an accurate representation of the real world” [203]. This definition implies that an important part of a VCT validation is careful consideration of the regulatory question to be answered. Validation of VCT models for imaging device efficacy have traditionally involved “fool the reader” type studies intended to characterize how difficult it is for an expert to subjectively distinguish a real image from a simulated one [204]. On the other hand, it has been suggested that validity and applicability of VCTs for regulatory use should hinge upon whether regulatory decisions based on the VCT are as good as, and more efficient than, decisions based upon data from currently used experimental methods [205]. Validation would thus involve input from the regulatory agency regarding how VCT data is used in the decision-making process. Alternatively, considering the effort involved in validation and potential applicability to multiple devices, VCT tools that incorporate a CHP could be evaluated as part of the Medical Device Development Tools program independent of a regulatory submission for a particular device [206]. This process could qualify a VCT platform for a specific context-of-use, potentially encompassing a range of regulatory questions or device types.

There are many regulatory questions that could be addressed more efficiently with CHPs, e.g., post-market quality control of clinical protocols could be a useful application of simulated images generated in a VCT due to the knowledge of ground truth [207]. Collecting and assessing validation evidence for CHPs, and simulations performed with them, is critical for their adoption in the regulatory process.

### III. CONCLUSION

If care is taken to proceed with the necessary rigor and caution, CHPs are an invaluable door opener for a wide range of computational life science applications that will benefit safety and efficacy assessments, including the development of novel therapies and medical devices, VCTs, and personalized medicine.

CHPs are becoming more detailed and realistic, more available and in greater number, and benefit from the rapid progress in simulation technology allowing increasingly complex simulations. This has opened new avenues in their use, beyond classic dosimetric applications, such as in medical device design and VCTs. For that purpose, it is important to achieve sufficient patient population coverage; a demand that is likely to be met through the development of morphing techniques, rather than continued *de novo* segmentation. Morphing techniques capable of adapting existing CHPs to information about an individual person (e.g., through medical image-registration) could broaden the use of detailed CHPs to patient-specific applications, such as personalized treatment planning and optimization, personalized dosimetry, and safety management for medical imaging. Hence, in parallel to the development of

simpler model creation approaches, e.g., using artificial intelligence techniques in medical image processing, it is expected that research efforts will shift from brute force segmentation to the facilitation of model parameterization and personalization through morphing techniques.

For biomedical imaging applications, CHPs allow experimentation which are too risky or impossible with real patients. CHPs allow complete control over the patient’s anatomy, the imaging device and any number of different imaging scenarios can be quickly and cost effectively investigated. CHPs provide a known truth that does not exist in real patients, providing a means with which to quantitatively evaluate, optimize, and compare different imaging system designs and configurations, data acquisition techniques, and image reconstruction and processing methods. As such, they are providing a valuable precursor or ultimately an alternative to HCTs.

Another important step will be the transition from developing CHPs as static geometrical representations of anatomy to incorporating information about dynamic behavior [16]. This includes, e.g., functionalization with nerve trajectories that have been enhanced with dynamic electrophysiological behavior [84], with breathing/heart-beat motion models, with vascular blood flow, or with thermoregulation models. Incorporating dynamics and physiology models in CHPs will broaden their application and facilitate sharing, compatibility, and interoperability of such physiology models, as well as their use in multiphysics modeling [208]. Recent developments use CHPs in multiscale simulations to estimate the radiobiological effect of clinical ionizing radiotherapy. This novel simulation technique combines multiscale MC simulations from organ to cell levels. At the cellular level, accumulated damages are computed using a spectrum-based accumulation algorithm and a predefined cellular damage database. Multiscale modeling is then performed under conventional fractionated irradiation [209]. Anatomical geometry and physiological dynamics can frequently not be treated as independent, therefore, integrated, functionalized CHPs are required.

However, the increasingly simple applicability of CHPs carries the risk of users incorporating such models in their simulations without awareness of the assumptions and limitations involved in creating such models. Even if the tissue delineation were perfect and the resolution infinitely detailed, as already the process of segmenting the complex and inhomogeneous tissue distribution into clearly distinct, locally homogeneous regions, is an approximation (unless the model is supplemented with tissue inhomogeneity maps). Approximations that are suitable for some applications are not acceptable for others, e.g., biomechanical applications require superior modeling of tendons and ligaments, as well as sliding tissue interfaces, which might not be necessary for most dosimetry applications [210]. Hence, the use of CHPs in applications that have not been previously validated, should typically be accompanied by careful consideration and discussion of the limitations, V&V, and uncertainty assessment. It is expected, that increasing model detailedness and fidelity, as well as improved statistics through an increasing number of independently created CHPs, will reduce the uncertainty and provide

information into the magnitude of certainty, e.g., randomly distributed modeling errors. However, such considerations will not provide insights into systematic errors that can only be analyzed through conscious investigation of the involved approximations and through experimental validation in a setup sufficiently relevant to the model context-of-use.

Finally, although the requirements for applications used by the different research communities (ionizing dosimetry, nonionizing EM radiation, passive car safety, etc.) in terms of segmented tissues and target heights, weights, and organ masses can be extremely variable, there is considerable overlap regarding techniques and tools under development for processing and modifying existing phantoms. While CHPs for ionizing and nonionizing radiation often have similar file formats, the models developed for car-crash safety contain the actual computational meshes and vendor-specific annotations. Nevertheless, sharing and reuse of segmentations created to cover specific target populations is likely to benefit all fields and could significantly reduce the effort needed to create new CHPs from image data.

#### DISCLAIMER

The opinions and/or conclusions expressed herein are solely those of the author and in no way imply a policy or position of the Food and Drug Administration. The mention of commercial products, their sources, or their use in connection with material reported herein is not to be construed as either an actual or implied endorsement of such products by the Department of Health and Human Services.

#### ACKNOWLEDGMENT

The authors would like to thank M. A. Cowdery and P. L. Bounds for the thorough and detailed proofread of this paper and the revision of 212 references.

#### REFERENCES

- [1] S. Agostinelli *et al.*, “Geant4—A simulation toolkit,” *Nucl. Instrum. Methods Phys. Res. A Accelerators Spectrometers Detectors Assoc. Equip.*, vol. 506, no. 3, pp. 250–303, Jul. 2003.
- [2] M. Leyton, *A Generative Theory of Shape*, vol. 2145. Berlin, Germany: Springer-Verlag, Nov. 2001.
- [3] I. Stroud, *Boundary Representation Modeling Techniques*. London, U.K.: Springer-Verlag, Jul. 2006.
- [4] M. Cristy and K. F. Eckerman, *Specific Absorbed Fractions of Energy at Various Ages From Internal Photon Sources, Part I: Methods*, Oak Ridge Nat. Lab., Oak Ridge, TN, USA, 1987.
- [5] X. G. Xu and K. F. Eckerman, *Handbook of Anatomical Models for Radiation Dosimetry*. Hoboken, NJ, USA: Taylor & Francis, 2009.
- [6] X. G. Xu, “An exponential growth of computational phantom research in radiation protection, imaging, and radiotherapy: A review of the fifty-year history,” *Phys. Med. Biol.*, vol. 59, no. 18, pp. R233–R302, 2014.
- [7] W. Schneider, T. Bortfeld, and W. Schlegel, “Correlation between CT numbers and tissue parameters needed for Monte Carlo simulations of clinical dose distributions,” *Phys. Med. Biol.*, vol. 45, no. 2, pp. 459–478, 2000.
- [8] J. S. Park *et al.*, “Visible Korean human: Improved serially sectioned images of the entire body,” *IEEE Trans. Med. Imag.*, vol. 24, no. 3, pp. 352–360, Mar. 2005.
- [9] Y. H. Na, B. Zhang, J. Zhang, P. F. Caracappa, and X. G. Xu, “Deformable adult human phantoms for radiation protection dosimetry: Anthropometric data representing size distributions of adult worker populations and software algorithms,” *Phys. Med. Biol.*, vol. 55, no. 13, pp. 3789–3811, 2010.

- [10] M. A. McDowell, C. D. Fryar, and C. L. Ogden, “Anthropometric reference data for children and adults: United States, 1988–1994,” *Vital Health Stat.*, vol. 249, pp. 1–68, Apr. 2009.
- [11] A. Christ *et al.*, “The virtual family—Development of surface-based anatomical models of two adults and two children for dosimetric simulations,” *Phys. Med. Biol.*, vol. 55, no. 2, pp. N23–N38, 2009.
- [12] M.-C. Gosselin *et al.*, “Development of a new generation of high-resolution anatomical models for medical device evaluation: The virtual population 3.0,” *Phys. Med. Biol.*, vol. 59, no. 18, pp. 5287–5303, 2014.
- [13] L. Kavan, S. M. Collins, J. Žára, and C. O’Sullivan, “Geometric skinning with approximate dual quaternion blending,” *ACM Trans. Graph.*, vol. 27, no. 4, pp. 1–23, 2008.
- [14] E. Cherubini, N. Chavannes, and N. Kuster, “Realistic skeleton based deformation of high-resolution anatomical human models for electromagnetic simulations,” in *Proc. 30th Annu. Meeting Bioelectromagn. Soc.*, San Diego, CA, USA, 2009, pp. 505–507.
- [15] T. C. F. Fonseca, R. Bogaerts, J. Hunt, and F. Vanhavere, “A methodology to develop computational phantoms with adjustable posture for WBC calibration,” *Phys. Med. Biol.*, vol. 59, no. 22, pp. 6811–6825, 2014.
- [16] B. Lloyd *et al.*, “Simulation and synthesis in medical imaging,” in *Proc. 1st Int. Workshop Simulat. Synth. Med. Imag. (SASHIMI)*, vol. 9968. Athens, Greece, Oct. 2016, pp. 13–22.
- [17] W. P. Segars, G. Sturgeon, S. Mendonca, J. Grimes, and B. M. W. Tsui, “4D XCAT phantom for multimodality imaging research,” *Med. Phys.*, vol. 37, no. 9, pp. 4902–4915, 2010.
- [18] W. P. Segars *et al.*, “Population of anatomically variable 4D XCAT adult phantoms for imaging research and optimization,” *Med. Phys.*, vol. 40, pp. 1–11, Apr. 2013.
- [19] W. P. Segars *et al.*, “The development of a population of 4D pediatric XCAT phantoms for imaging research and optimization,” *Med. Phys.*, vol. 42, no. 8, pp. 4719–4726, 2015.
- [20] A. Akbarzadeh *et al.*, “Evaluation of whole-body MR to CT deformable image registration,” *J. Appl. Clin. Med. Phys.*, vol. 14, no. 4, pp. 238–253, 2013.
- [21] D. Ali-Hamadi *et al.*, “Anatomy transfer,” *ACM Trans. Graph.*, vol. 32, no. 6, pp. 1–8, 2013.
- [22] D. Xiao *et al.*, “An accurate 3D shape context based non-rigid registration method for mouse whole-body skeleton registration,” in *Proc. Med. Imag. Process.*, vol. 7962. Orlando, FL, USA, 2011, Art. no. 79620V.
- [23] H. Wang, D. B. Stout, and A. F. Chatziioannou, “Estimation of mouse organ locations through registration of a statistical mouse atlas with micro-CT images,” *IEEE Trans. Med. Imag.*, vol. 31, no. 1, pp. 88–102, Jan. 2012.
- [24] J. Zhang, X. G. Xu, C. Shi, and M. Fuss, “Development of a geometry-based respiratory motion-simulating patient model for radiation treatment dosimetry,” *J. Appl. Clin. Med. Phys.*, vol. 9, no. 1, pp. 16–28, 2008.
- [25] M. C. Han *et al.*, “Continuously deforming 4D voxel phantom for realistic representation of respiratory motion in Monte Carlo dose calculation,” *IEEE Trans. Nucl. Sci.*, vol. 63, no. 6, pp. 2918–2924, Dec. 2016.
- [26] J. Eom, X. G. Xu, S. De, and C. Shi, “Predictive modeling of lung motion over the entire respiratory cycle using measured pressure-volume data, 4DCT images, and finite-element analysis,” *Med. Phys.*, vol. 37, no. 8, pp. 4389–4400, 2010.
- [27] W. P. Segars, A. I. Veress, G. M. Sturgeon, and E. Samei, “Incorporation of the living heart model into the 4D XCAT phantom for cardiac imaging research,” *IEEE Trans. Radiat. Plasma Med. Sci.*, to be published, doi: [10.1109/TRPMS.2018.2823060](https://doi.org/10.1109/TRPMS.2018.2823060).
- [28] C. H. Kim, J. H. Jeong, W. E. Boch, K.-W. Cho, and S. B. Hwang, “A polygon-surface reference Korean male phantom (PSRK-man) and its direct implementation in Geant4 Monte Carlo simulation,” *Phys. Med. Biol.*, vol. 56, no. 10, pp. 3137–3161, 2011.
- [29] Y. S. Yeom, J. H. Jeong, M. C. Han, and C. H. Kim, “Tetrahedral-mesh-based computational human phantom for fast Monte Carlo dose calculations,” *Phys. Med. Biol.*, vol. 59, no. 12, pp. 3173–3185, 2014.
- [30] International Commission on Radiological Protection, “Adult reference computational phantoms, Publication 110,” *Ann. ICRP*, vol. 39, no. 2, pp. 1–164, Apr. 2009.
- [31] C. H. Kim *et al.*, “New mesh-type phantoms and their dosimetric applications, including emergencies,” *Ann. ICRP*, vol. 47, nos. 3–4, pp. 45–62, 2018.
- [32] M. C. Han, Y. S. Yeom, C. H. Kim, S. Kim, and J. W. Sohn, “New approach based on tetrahedral-mesh geometry for accurate 4D Monte Carlo patient-dose calculation,” *Phys. Med. Biol.*, vol. 60, no. 4, pp. 1601–1612, 2015.

- [33] P. Pathmanathan, R. A. Gray, V. J. Romero, and T. M. Morrison, "Applicability analysis of validation evidence for biomedical computational models," *J. Verification Validation Uncertainty Quantification*, vol. 2, no. 2, pp. 1–11, 2017.
- [34] ASME V&V 40. *Assessing Credibility of Computational Models Through Verification and Validation: Application to Medical Devices*. Accessed: Dec. 8, 2018. [Online]. Available: <http://go.asme.org/VnV40Committee>
- [35] S. Farcito, B. Lloyd, and N. Kuster, *Verification Report VIP001AA201704: Virtual Population Models V3.X Verification of Anatomy, Physiology and Segmentation/Generation Procedures*, IT'IS Found., Zürich, Switzerland, 2017.
- [36] R. McGurk *et al.*, "Extension of the NCAT phantom for the investigation of intra-fraction respiratory motion in IMRT using 4D Monte Carlo," *Phys. Med. Biol.*, vol. 55, no. 5, pp. 1475–1490, 2010.
- [37] M. Serban, E. Heath, G. Stroian, D. L. Collins, and J. Seuntjens, "A deformable phantom for 4D radiotherapy verification: Design and image registration evaluation," *Med. Phys.*, vol. 35, no. 3, pp. 1094–1102, Mar. 2008.
- [38] S. K. Warfield, K. H. Zou, and W. M. Wells, "Simultaneous truth and performance level estimation (STAPLE): An algorithm for the validation of image segmentation," *IEEE Trans. Med. Imag.*, vol. 23, no. 7, pp. 903–921, Jul. 2004.
- [39] B. H. Menze *et al.*, "The multimodal brain tumor image segmentation benchmark (BRATS)," *IEEE Trans. Med. Imag.*, vol. 34, no. 10, pp. 1993–2024, Oct. 2015.
- [40] M. I. Iacono *et al.*, "MIDA: A multimodal imaging-based detailed anatomical model of the human head and neck," *PLoS ONE*, vol. 10, Apr. 2015, Art. no. e0124126.
- [41] P. Hariharan *et al.*, "Use of the FDA nozzle model to illustrate validation techniques in computational fluid dynamics (CFD) simulations," *PLoS ONE*, vol. 12, no. 6, 2017, Art. no. e0178749.
- [42] H. H. Pennes, "Analysis of tissue and arterial blood temperatures in the resting human forearm," *J. Appl. Phys.*, vol. 1, no. 2, pp. 93–122, 1948.
- [43] M. Murbach *et al.*, "Thermal tissue damage model analyzed for different whole-body SAR and scan durations for standard MR body coils," *Magn. Resonance Med.*, vol. 71, no. 1, pp. 421–431, 2014.
- [44] C. Guiot *et al.*, "Perfusion and thermal field during hyperthermia. Experimental measurements and modelling in recurrent breast cancer," *Phys. Med. Biol.*, vol. 43, no. 10, pp. 2831–2843, 1998.
- [45] Open Ergonomics. *PeopleSize 2008 of Customers to Come—Visual Anthropometry Software*. Accessed: Dec. 8, 2018. [Online]. Available: <http://www.openeng.com/ps/index.html>
- [46] D. Broggio *et al.*, "Construction of an extended library of adult male 3D models: Rationale and results," *Phys. Med. Biol.*, vol. 56, no. 23, pp. 7659–7662, 2011.
- [47] W. S. Snyder, M. R. Ford, and G. G. Warner, *Estimates of Specific Absorbed Fractions for Photon Sources Uniformly Distributed in Various Organs of a Heterogeneous Phantom*, Soc. Nucl. Med., New York, NY, USA, 1978.
- [48] IT'IS. *ViP2 IT'IS Foundation*. Accessed: Dec. 8, 2018. [Online]. Available: <https://itis.swiss/virtual-population/virtual-population/vip2/>
- [49] T. Samaras, P. Regli, and N. Kuster, "Electromagnetic and heat transfer computations for non-ionizing radiation dosimetry," *Phys. Med. Biol.*, vol. 45, no. 8, pp. 2233–2246, 2000.
- [50] A. Kyriakou *et al.*, "A review of numerical and experimental compensation techniques for skull-induced phase aberrations in transcranial focused ultrasound," *Int. J. Hyperthermia*, vol. 30, no. 1, pp. 36–46, Feb. 2014.
- [51] W. P. Segars, M. Mahesh, T. J. Beck, E. C. Frey, and B. M. W. Tsui, "Realistic CT simulation using the 4D XCAT phantom," *Med. Phys.*, vol. 35, no. 8, pp. 3800–3808, 2008.
- [52] J. Michaelson, J. Forman, R. Kent, and S. Kuppa, "Rear seat occupant safety: Kinematics and injury of PMHS restrained by a standard 3-point belt in frontal crashes," *Stapp Car Crash J.*, vol. 52, pp. 295–325, Nov. 2008.
- [53] International Commission on Radiological Protection, "Reference man: Anatomical, physiological and metabolic characteristics, Publication 23," *Ann. ICRP*, vol. 4, nos. 3–4, 1974.
- [54] International Commission on Radiological Protection, "Basic anatomical and physiological data for use in radiological protection reference values, Publication 89," *Ann. ICRP*, vol. 32, nos. 3–4, p. 265, 2002.
- [55] P. W. Nicholson, "Specific impedance of cerebral white matter," *Exp. Neurol.*, vol. 13, no. 4, pp. 386–401, 1965.
- [56] C. Gabriel, S. Gabriel, and E. Corthout, "The dielectric properties of biological tissues: I. Literature survey," *Phys. Med. Biol.*, vol. 41, no. 11, pp. 2231–2249, 1996.
- [57] P. R. Bakic, K. J. Myers, S. J. Glick, and A. D. A. Maidment, "Virtual tools for the evaluation of breast imaging: State-of-the science and future directions," in *Proc. Int. Workshop Breast Imag.*, Malmö, Sweden, 2016, pp. 518–524.
- [58] F. Duck, *Physical Properties of Tissues*, 1st ed. San Diego, CA, USA: Academic, 1990.
- [59] F. Kreith, *The CRC Handbook of Mechanical Engineering*, 2nd ed. Hoboken, NJ, USA: CRC Press, 2004.
- [60] IT'IS. *Database of Tissue Properties*. Accessed: Dec. 8, 2018. [Online]. Available: <https://www.itis.ethz.ch/database>
- [61] P. A. Havgall, E. Neufeld, M. C. Gosselin, A. Klingenbock, and N. Kuster, *IT'IS Database for Thermal and Electromagnetic Parameters of Biological Tissues*, IT'IS Found., Zürich, Switzerland, 2017.
- [62] R. L. McIntosh and V. Anderson, "A comprehensive tissue properties database provided for the thermal assessment of a human at rest," *Biophys. Rev. Lett.*, vol. 5, no. 3, pp. 129–151, 2010.
- [63] S. Gabriel, R. W. Lau, and C. Gabriel, "The dielectric properties of biological tissues: II. Measurements in the frequency range 10 Hz to 20 GHz," *Phys. Med. Biol.*, vol. 41, no. 11, pp. 2251–2269, 1996.
- [64] S. Gabriel, R. W. Lau, and C. Gabriel, "The dielectric properties of biological tissues: III. Parametric models for the dielectric spectrum of tissues," *Phys. Med. Biol.*, vol. 41, no. 11, pp. 2271–2293, 1996.
- [65] K. S. Cole and R. H. Cole, "Dispersion and absorption in dielectrics I. Alternating current characteristics," *J. Chem. Phys.*, vol. 9, no. 4, pp. 341–351, 1941.
- [66] IFAC, *An Internet Resource for the Calculation of the Dielectric Properties of Body Tissues*, FAC-CNR, Florence, Italy, 2015.
- [67] T. Wu, T. S. Rappaport, and C. M. Collins, "Safe for generations to come: Considerations of safety for millimeter waves in wireless communications," *IEEE Microw. Mag.*, vol. 16, no. 2, pp. 65–84, Mar. 2015.
- [68] T. Wu, T. S. Rappaport, and C. M. Collins, "The human body and millimeter-wave wireless communication systems: Interactions and implications," in *Proc. IEEE Int. Conf. Commun. (ICC)*, London, U.K., 2015, pp. 2423–2429.
- [69] E. Neufeld, M. M. Paulides, G. C. V. Rhoon, and N. Kuster, "Numerical modeling for simulation and treatment planning of thermal therapy," in *Imaging in Medical Diagnosis and Therapy*. Boca Raton, FL, USA: Taylor & Francis, 2013, pp. 119–138.
- [70] H. Arkin, L. X. Xu, and K. R. Holmes, "Recent developments in modeling heat transfer in blood perfused tissues," *IEEE Trans. Biomed. Eng.*, vol. 41, no. 2, pp. 97–107, Feb. 1994.
- [71] A. Kyriakou, "Multi-physics computational modeling of focused ultrasound therapies," Ph.D. dissertation, ETH Zürich, Zürich, Switzerland, 2015.
- [72] T. D. Mast, "Empirical relationships between acoustic parameters in human soft tissues," *Acoust. Res. Lett. Online*, vol. 1, pp. 37–42, Nov. 2000.
- [73] U. Vyas *et al.*, "Non-invasive patient-specific acoustic property estimation for MR-guided focused ultrasound surgery," in *Proc. AIP Conf.*, vol. 1481, 2012, pp. 419–425.
- [74] H. Q. Woodard and D. R. White, "The composition of body tissues," *Brit. J. Radiol.*, vol. 59, no. 708, pp. 1209–1218, 1986.
- [75] Y. C. Fung, "Bio-viscoelastic solids," in *Biomechanics*. New York, NY, USA: Springer, 1981, pp. 196–260. [Online]. Available: [https://doi.org/10.1007/978-1-4757-1752-5\\_7](https://doi.org/10.1007/978-1-4757-1752-5_7)
- [76] G. Low, S. A. Kruse, and D. J. Lomas, "General review of magnetic resonance elastography," *World J. Rad.*, vol. 8, no. 1, pp. 59–72, Jan. 2016.
- [77] X. G. Xu, V. Taranenko, J. Zhang, and C. Shi, "A boundary-representation method for designing whole-body radiation dosimetry models: Pregnant females at the ends of three gestational periods—RP1-P3, -P6 and -P9," *Phys. Med. Biol.*, vol. 52, no. 23, pp. 7023–7044, Dec. 2007.
- [78] IT'IS. *Virtual Population V3.X*. Accessed: Dec. 8, 2018. [Online]. Available: <https://www.itis.ethz.ch/vip3>
- [79] L. Liu *et al.*, "An ICRP-based Chinese adult male voxel model and its absorbed dose for idealized photon exposures—The skeleton," *Phys. Med. Biol.*, vol. 54, no. 21, pp. 6675–6690, 2009.

- [80] T. Nagaoka *et al.*, "Development of realistic high-resolution whole-body voxel models of Japanese adult males and females of average height and weight, and application of models to radio-frequency electromagnetic-field dosimetry," *Phys. Med. Biol.*, vol. 49, no. 1, pp. 1–15, 2004.
- [81] T. Xie, N. Kuster, and H. Zaidi, "Computational hybrid anthropometric paediatric phantom library for internal radiation dosimetry," *Phys. Med. Biol.*, vol. 62, no. 8, pp. 3263–3283, 2017.
- [82] H. Zaidi and X. G. Xu, "Computational anthropomorphic models of the human anatomy: The path to realistic Monte Carlo modeling in radiological sciences," *Annu. Rev. Biomed. Eng.*, vol. 9, pp. 471–500, Feb. 2007.
- [83] H. Zaidi and B. M. W. Tsui, "Review of computational anthropomorphic anatomical and physiological models," *Proc. IEEE*, vol. 97, no. 12, pp. 1938–1953, Dec. 2009.
- [84] E. Neufeld, A. M. Cassarà, H. Montanaro, N. Kuster, and W. Kainz, "Functionalized anatomical models for EM-neuron interaction modeling," *Phys. Med. Biol.*, vol. 61, no. 12, pp. 4390–4401, Jun. 2016.
- [85] E. Abadi *et al.*, "Modeling lung architecture in the XCAT series of phantoms: Physiologically based airways, arteries and veins," *IEEE Trans. Med. Imag.*, vol. 37, no. 3, pp. 693–702, Mar. 2018.
- [86] E. Abadi, B. Harrawood, A. Kapadia, W. P. Segars, and E. and Samei, "Development of a fast, voxel-based, and scanner-specific CT simulator for image-quality-based virtual clinical trials," in *Proc. SPIE Med. Imag. Phys. Med. Imag.*, Houston, TX, USA, 2018, Art. no. 105733U.
- [87] E. Abadi *et al.*, "Airways, vasculature, and interstitial tissue: Anatomically informed computational modeling of human lungs for virtual clinical trials," in *Proc. Vol. 10132 Med. Imag. Phys. Med. Imag.*, Orlando, FL, USA, 2017, Art. no. 101321Q.
- [88] P. Sahbaee, W. P. Segars, D. Marin, R. C. Nelson, and E. Samei, "The effect of contrast material on radiation dose at CT: Part I. Incorporation of contrast material dynamics in anthropomorphic phantoms," *Radiology*, vol. 283, no. 3, pp. 739–748, 2017.
- [89] W. P. Segars *et al.*, "Application of the 4-D XCAT phantoms in biomedical imaging and beyond," *IEEE Trans. Med. Imag.*, vol. 37, no. 3, pp. 680–692, Mar. 2018.
- [90] N. Petoussi-Hens, M. Zanki, U. Fill, and D. Regulla, "The GSF family of voxel phantoms," *Phys. Med. Biol.*, vol. 47, no. 1, pp. 89–106, 2002.
- [91] X. G. Xu, T. C. Chao, and A. Bozkurt, "VIP-Man: An image-based whole-body adult male model constructed from color photographs of the visible human project for multi-particle Monte Carlo calculations," *Health Phys.*, vol. 78, no. 5, pp. 476–486, May 2000.
- [92] M. J. Ackerman, "The visible human project," *Proc. IEEE*, vol. 86, no. 3, pp. 504–511, Mar. 1998.
- [93] C. Lee, J. L. Williams, C. Lee, and W. E. Bolch, "The UF series of tomographic computational phantoms of pediatric patients," *Med. Phys.*, vol. 32, no. 12, pp. 3537–3548, Dec. 2005.
- [94] C. Lee *et al.*, "Hybrid computational phantoms of the male and female newborn patient: NURBS-based whole-body models," *Phys. Med. Biol.*, vol. 52, no. 12, pp. 3309–3333, 2007.
- [95] C. Lee *et al.*, "The UF family of reference hybrid phantoms for computational radiation dosimetry," *Phys. Med. Biol.*, vol. 55, no. 2, pp. 339–363, 2010.
- [96] C. H. Kim, S. H. Choi, J. H. Jeong, C. Lee, and M. S. Chung, "HDRK-Man: A whole-body voxel model based on high-resolution color slice images of a Korean adult male cadaver," *Phys. Med. Biol.*, vol. 53, no. 15, pp. 4093–4106, 2008.
- [97] Y. S. Yeom *et al.*, "HDRK-Woman: Whole-body voxel model based on high-resolution color slice images of Korean adult female cadaver," *Phys. Med. Biol.*, vol. 59, no. 14, pp. 3969–3984, 2014.
- [98] J. W. Massey and A. E. Yilmaz, "AustinMan and AustinWoman: High-fidelity, anatomical voxel models developed from the VHP color images," in *Proc. 38th Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. (EMBC)*, Orlando, FL, USA, 2016, pp. 3346–3349.
- [99] Creative Commons. *Attribution-NonCommercial-NoDerivs 3.0 Unported—CC BY-NC-ND 3.0*. Accessed: Dec. 8, 2018. [Online]. Available: <https://creativecommons.org/licenses/by-nc-nd/3.0/>
- [100] T. Janak. (2017). *PIPER FEM Child Model*. [Online]. Available: <https://gitlab.inria.fr/piper/child>
- [101] M. Svensson, K. Brolin, and A. Linder. (2018). *Virtual Vehicle-Safety Assessment: Open Source Human Body Models Addressing Gender Diversity (ViVA I)*. [Online]. Available: <https://www.chalmers.se/en/projects/Pages/ViVA—Virtual-Vehicle-safety-Assessment—Open-Source-Human-Body-Models-addressing-gender-diversity.aspx>
- [102] GHBMC. (2017). *Global Human Body Model Consortium*. [Online]. Available: <https://www.ghbmc.com/>
- [103] E. Jolivet, Y. Lafon, P. Petit, and P. Beillas, "Comparison of kriging and moving least square methods to change the geometry of human body models," *Stapp Car Crash J.*, vol. 59, pp. 337–357, Nov. 2015.
- [104] S. L. Schoell, A. A. Weaver, N. A. Vavalle, and J. D. Stitzel, "Age- and sex-specific thorax finite element model development and simulation," *Traffic Injury Prevent.*, vol. 16, no. S1, pp. S57–S65, 2015.
- [105] D. Schwartz, B. Guleyupoglu, B. Koya, J. D. Stitzel, and F. S. Gayzik, "Development of a computationally efficient full human body finite element model," *Traffic Injury Prevent.*, vol. 16, no. S1, pp. S49–S56, 2015.
- [106] Livermore Software Technology Corporation. (2018). *Total Human Model for Safety—THUMS*. [Online]. Available: <http://www.lstc.com/thums>
- [107] E. Hwang *et al.*, *Rapid Development of Diverse Human Body Models for Crash Simulations Through Mesh Morphing*, SAE Mobilus, Warrendale, PA, USA, 2016.
- [108] W. E. Bolch, C. Lee, M. Wayson, and P. Johnson, "Hybrid computational phantoms for medical dose reconstruction," *Radiat. Environ. Biophys.*, vol. 49, no. 2, pp. 155–168, 2010.
- [109] A. Ding, M. M. Mille, T. Liu, P. F. Caracappa, and X. G. Xu, "Extension of RPI-adult male and female computational phantoms to obese patients and a Monte Carlo study of the effect on CT imaging dose," *Phys. Med. Biol.*, vol. 57, no. 9, pp. 2441–2459, 2012.
- [110] A. Ding *et al.*, "VirtualDose: A software for reporting organ doses from CT for adult and pediatric patients," *Phys. Med. Biol.*, vol. 60, no. 14, pp. 5601–5625, 2015.
- [111] E. J. Stepusin, D. J. Long, E. L. Marshall, and W. E. Bolch, "Assessment of different patient-to-phantom matching criteria applied in Monte Carlo-based computed tomography dosimetry," *Med. Phys.*, vol. 44, no. 10, pp. 5498–5508, 2017.
- [112] A. M. Geyer, S. O'Reilly, C. Lee, D. J. Long, and W. E. Bolch, "The UF/NCI family of hybrid computational phantoms representing the current U.S. population of male and female children, adolescents, and adults—Application to CT dosimetry," *Phys. Med. Biol.*, vol. 59, no. 18, pp. 5225–5242, 2014.
- [113] W. Fu *et al.*, "Breast dose reduction with organ-based, wide-angle tube current modulated CT," *J. Med. Imag.*, vol. 4, no. 3, 2017, Art. no. 0321208.
- [114] D. Borrego, D. A. Siragusa, S. Balter, and W. E. Bolch, "A hybrid phantom system for patient skin and organ dosimetry in fluoroscopically guided interventions," *Med. Phys.*, vol. 44, no. 9, pp. 4928–4942, 2017.
- [115] P. B. Johnson *et al.*, "Skin dose mapping for fluoroscopically guided interventions," *Med. Phys.*, vol. 38, no. 10, pp. 5490–5499, 2011.
- [116] A. Zvereva *et al.*, "Feasibility of reducing differences in estimated doses in nuclear medicine between a patient-specific and a reference phantom," *Physica Medica*, vol. 39, pp. 100–112, Jul. 2017.
- [117] C. Kim *et al.*, "The reference phantoms: Voxel versus polygon," *Ann. ICRP*, vol. 45, pp. 188–201, Jun. 2016.
- [118] H. Petroccia *et al.*, "A hybrid phantom Monte Carlo-based method for historical reconstruction of organ doses in patients treated with cobalt-60 for Hodgkin's lymphoma," *Phys. Med. Biol.*, vol. 62, no. 15, pp. 6261–6289, 2017.
- [119] P. Papadimitriou *et al.*, "A personalized, Monte Carlo-based method for internal dosimetric evaluation of radiopharmaceuticals in children," *Med. Phys.*, vol. 45, no. 8, pp. 3939–3949, 2018.
- [120] S. Senthamizchelvan *et al.*, "Tumor dosimetry and response for <sup>153</sup>Sm-ethylenediamine tetramethylene phosphonic acid therapy of high-risk osteosarcoma," *J. Nucl. Med.*, vol. 53, no. 2, pp. 215–224, 2012.
- [121] M. G. Stabin, R. B. Sparks, and E. Crowe, "OLINDA/EXM: The second-generation personal computer software for internal dose assessment in nuclear medicine," *J. Nucl. Med.*, vol. 46, no. 6, pp. 1023–1027, 2005.
- [122] J. M. Osepchuk and R. C. Petersen, "Safety standards for exposure to RF electromagnetic fields," *IEEE Microw. Mag.*, vol. 2, no. 2, pp. 57–69, Jun. 2001.
- [123] C. H. Durney, M. F. Iskander, H. Massoudi, and C. C. Johnson, "An empirical formula for broad-band SAR calculations of prolate spheroidal models of humans and animals," *IEEE Trans. Microw. Theory Techn.*, vol. 27, no. 8, pp. 758–763, Aug. 1979.
- [124] H. P. Schwan and K. Li, "Hazards due to total body irradiation by radar," *Proc. IRE*, vol. 44, no. 11, pp. 1572–1581, Nov. 1956.
- [125] N. Kuster and Q. Balzano, "Energy absorption mechanism by biological bodies in the near field of dipole antennas above 300 MHz," *IEEE Trans. Veh. Technol.*, vol. 41, no. 1, pp. 17–23, Feb. 1992.

- [126] International Commission on Non-Ionizing Radiation Protection, "Guidelines for limiting exposure to time-varying electric, magnetic, and electromagnetic fields (up to 300 GHz). International commission on non-ionizing radiation protection," *Health Phys.*, vol. 74, no. 4, pp. 494–522, 1998.
- [127] *Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3kHz to 300GHz*, IEEE Standard C95.1:2005, 2005.
- [128] *Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Head From Wireless Communications Devices: Measurement Techniques*, IEEE Standard 1528:2013, 2013.
- [129] *Human Exposure to Radio Frequency Fields From Hand Held and Body Mounted Wireless Communication Devices—Human Models, Instrumentation and Procedures, Part 1: Devices Used Next to the Ear (Frequency Range of 300 MHz to 6 GHz)*, IEC Standard 62209-1:2016, 2016.
- [130] *Human Exposure to Radio Frequency Fields From Hand-Held and Body-Mounted Wireless Communication Devices—Human Models, Instrumentation, and Procedures, Part 2: Procedure to Determine the Specific Absorption Rate (SAR) for Wireless Communication Devices Used in Close Proximity to the Human Body (Frequency Range of 30 MHz to 6 GHz)*, IEC Standard 62209-2:2010, 2010.
- [131] P. J. Dimbylow and S. M. Mann, "SAR calculations in an anatomically realistic model of the head for mobile communication transceivers at 900 MHz and 1.8 GHz," *Phys. Med. Biol.*, vol. 39, no. 10, pp. 1537–1553, 1994.
- [132] A. Drossos, V. Santomaa, and N. Kuster, "The dependence of electromagnetic energy absorption upon human head tissue composition in the frequency range of 300–3000 MHz," *IEEE Trans. Microw. Theory Techn.*, vol. 48, no. 11, pp. 1988–1995, Nov. 2000.
- [133] A. Christ, T. Samaras, A. Klingenböck, and N. Kuster, "Characterization of the electromagnetic near-field absorption in layered biological tissue in the frequency range from 30 MHz to 6,000 MHz," *Phys. Med. Biol.*, vol. 51, no. 19, pp. 4951–4965, 2006.
- [134] F. Schönborn, M. Burkhardt, and N. Kuster, "The difference of EM energy absorption between adults and children," *Health Phys.*, vol. 74, no. 2, pp. 160–168, 1998.
- [135] M. Burkhardt and N. Kuster, "Appropriate modeling of the ear for compliance testing of handheld MTE with SAR safety limits at 900/1800 MHz," *IEEE Trans. Microw. Theory Techn.*, vol. 48, no. 11, pp. 1927–1934, Nov. 2000.
- [136] B. B. Beard *et al.*, "Comparisons of computed mobile phone induced SAR in the SAM phantom to that in anatomically correct models of the human head," *IEEE Trans. Electromagn. Compat.*, vol. 48, no. 2, pp. 397–407, May 2006.
- [137] W. Kainz *et al.*, "Dosimetric comparison of the specific anthropomorphic mannequin (SAM) to 14 anatomical head models using a novel definition for the mobile phone positioning," *Phys. Med. Biol.*, vol. 50, no. 14, pp. 3423–3445, 2005.
- [138] J. Keshvari, M. Kivento, A. Christ, and G. Bit-Babik, "Large scale study on the variation of RF energy absorption in the head and brain regions of adults and children and evaluation of the SAM phantom conservativeness," *Phys. Med. Biol.*, vol. 61, no. 8, pp. 2991–3008, 2016.
- [139] *Determination of RF Field Strength, Power Density and SAR in the Vicinity of Radiocommunication Base Stations for the Purpose of Evaluating Human Exposure*, IEC Standard 62232:2017, 2017.
- [140] M.-C. Gosselin *et al.*, "Estimation formulas for the specific absorption rate in humans exposed to base-station antennas," *IEEE Trans. Electromagn. Compat.*, vol. 53, no. 4, pp. 909–922, Nov. 2011.
- [141] A. Christ, M. Douglas, J. Nadakuduti, and N. Kuster, "Assessing human exposure to electromagnetic fields from wireless power transmission systems," *Proc. IEEE*, vol. 101, no. 6, pp. 1482–1493, Jun. 2013.
- [142] A. Christ *et al.*, "Exposure of the human body to professional and domestic induction cooktops compared to the basic restrictions," *Bioelectromagnetics*, vol. 33, no. 8, pp. 695–705, 2012.
- [143] *Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body From Wireless Communications Devices, 30 MHz to 6 GHz, Part 1: General Requirements for Using the Finite Difference Time-Domain (FDTD) Method for SAR Calculations*, IEC/IEEE Standard 62704-1:2017, 2017.
- [144] *Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body From Wireless Communications Devices, 30 MHz to 6 GHz, Part 3: Specific Requirements for Using the Finite Difference Time Domain (FDTD) Method for SAR Calculations of Mobile Phones*, IEC/IEEE Standard 62704-3:2017, 2017.
- [145] *Particular Requirements for the Safety of Magnetic Resonance Equipment, 3rd Edition*, IEC Standard 60601-2-33+AMD2:2015, 2015.
- [146] E. Neufeld *et al.*, "Analysis of the local worst-case SAR exposure caused by an MRI multi-transmit body coil in anatomical models of the human body," *Phys. Med. Biol.*, vol. 56, no. 15, pp. 4649–4659, 2011.
- [147] M. Murbach *et al.*, "Virtual population-based assessment of the impact of 3 Tesla radiofrequency shimming and thermoregulation on safety and B1 + uniformity," *Mag. Reson. Med.*, vol. 76, no. 3, pp. 986–997, 2016.
- [148] *Assessment of the Safety of Magnetic Resonance Imaging for Patients With an Active Implantable Medical Device*, ISO Standard TS 10974:2018, 2018.
- [149] E. Neufeld, S. Kuehn, G. Szekely, and N. Kuster, "Measurement, simulation and uncertainty assessment of implant heating during MRI," *Phys. Med. Biol.*, vol. 54, no. 13, pp. 4151–4169, 2009.
- [150] *Standard Test Method for Measurement of Radio Frequency Induced Heating Near Passive Implants During Magnetic Resonance Imaging*, ASTM Standard F2182-02:2002, 2002.
- [151] S. I. Alekseev and M. C. Ziskin, "Human skin permittivity determined by millimeter wave reflection measurements," *Bioelectromagnetics*, vol. 28, no. 5, pp. 331–339, 2007.
- [152] M. C. Ziskin, S. I. Alekseev, K. R. Foster, and Q. Balzano, "Tissue models for RF exposure evaluation at frequencies above 6GHz," *Bioelectromagnetics*, vol. 39, no. 3, pp. 173–189, 2018.
- [153] "Quantities and units in radiation protection dosimetry," Int. Commission Radiat. Units Meas., Bethesda, MD, USA, Rep. 51, 1993.
- [154] International Commission on Radiological Protection, "The 2007 recommendations of the international commission on radiological protection, Publication 103," *Ann. ICRP*, vol. 37, nos. 2–4, pp. 1–332, 2007.
- [155] H. L. Fisher and W. S. Snyder, "Distribution of dose in the body from a source of gamma rays distributed uniformly in an organ," in *Proc. 1st Int. Congr. Radiat. Protect.*, Oxford, U.K., 1968, pp. 1473–1486.
- [156] M. Cristy, *Mathematical Phantoms Representing Children of Various Ages for Use in Estimates of Internal Dose*, Oak Ridge Nat. Lab., Oak Ridge, TN, USA, 1980.
- [157] J. Becker, M. Zankl, U. Fill, and C. Hoeschen, "Katja—The 24<sup>th</sup> week of virtual pregnancy for dosimetric calculations," *Polish J. Med. Phys. Eng.*, vol. 14, no. 1, pp. 13–19, 2008.
- [158] M. Caon, "Voxel-based computational models of real human anatomy: A review," *Radiat. Environ. Biophys.*, vol. 42, no. 4, pp. 229–235, 2004.
- [159] M. Caon, G. Bibbo, and J. Pattison, "An EGS4-ready tomographic computational model of a fourteen year-old female torso for calculating organ doses from CT examinations," *Phys. Med. Biol.*, vol. 44, no. 9, pp. 2213–2225, 1999.
- [160] P. J. Dimbylow, "Development of the female voxel phantom, NAOMI, and its application to calculations of induced current densities and electric fields from applied low frequency magnetic and electric fields," *Phys. Med. Biol.*, vol. 50, no. 6, pp. 1047–1070, 2005.
- [161] P. J. Dimbylow, "The development of realistic voxel phantoms for electromagnetic field dosimetry," in *Proc. Workshop Voxel Phantom Develop.*, Chilton, U.K., 1996, pp. 1–7.
- [162] U. Fill, M. Zankl, N. Petoussi-Henss, M. Siebert, and D. Regulla, "Adult female voxel models of different stature and photon conversion coefficients for radiation protection," *Health Phys.*, vol. 86, no. 3, pp. 253–272, 2004.
- [163] R. Kramer, H. J. Khoury, J. W. Vieira, and V. J. M. Lima, "MAX06 and FAX06: Update of two adult human phantoms for radiation protection dosimetry," *Phys. Med. Biol.*, vol. 51, no. 14, pp. 3331–3346, 2006.
- [164] R. Kramer *et al.*, "All about FAX: A female adult voxel phantom for Monte Carlo calculations in radiation protection dosimetry," *Phys. Med. Biol.*, vol. 49, no. 23, pp. 5203–5216, 2004.
- [165] R. Kramer, J. W. Vieira, H. J. Khoury, F. R. A. Lima, and D. Fuele, "All about MAX: A male adult voxel phantom for Monte Carlo calculations in radiation protection dosimetry," *Phys. Med. Biol.*, vol. 48, no. 10, pp. 1239–1262, 2003.
- [166] M. Zankl *et al.*, "The construction of computer tomographic phantoms and their application in radiology and radiation protection," *Radiat. Environ. Biophys.*, vol. 27, no. 2, pp. 153–164, 1988.
- [167] M. Zankl and A. Wittmann, "The adult male voxel model 'Golem' segmented from whole body CT patient data," *Radiat. Environ. Biophys.*, vol. 40, no. 2, pp. 153–162, 2001.
- [168] D. G. Jones, "A realistic anthropomorphic phantom for calculating organ doses arising from external photon irradiation," *Radiat. Protect. Dosimetry*, vol. 72, no. 1, pp. 21–29, 1997.

- [169] T. C. Chao, A. Bozkurt, and X. G. Xu, "Conversion coefficients based on the VIP-Man anatomical model and EGS4-VLSI code for external monoenergetic photons from 10 keV to 10 MeV," *Health Phys.*, vol. 81, no. 2, pp. 163–183, 2001.
- [170] M. Zankl, U. Fill, N. Petoussi-Henss, and D. Regulla, "Organ dose conversion coefficients for external photon irradiation of male and female voxel models," *Phys. Med. Biol.*, vol. 47, no. 14, pp. 2367–2385, 2002.
- [171] R. Kramer, J. W. Vieira, H. J. Khoury, and F. L. de Andreda, "MAX meets ADAM: A dosimetric comparison between a voxel-based and a mathematical model for external exposure to photons," *Phys. Med. Biol.*, vol. 49, no. 6, pp. 887–910, 2004.
- [172] H. Schlattl, M. Zankl, and N. Petoussi-Henss, "Organ dose conversion coefficients for voxel models of the reference male and female from idealized photon exposures," *Phys. Med. Biol.*, vol. 52, no. 8, pp. 2123–2145, 2007.
- [173] D. G. Jones, "A realistic anthropomorphic phantom for calculating specific absorbed fractions of energy deposited from internal gamma emitters," *Radiat. Protect. Dosimetry*, vol. 79, nos. 1–4, pp. 411–414, 1998.
- [174] T. Smith, N. Petoussi-Henss, and M. Zankl, "Comparison of internal radiation doses estimated by MIRD and voxel techniques for a 'family' of phantoms," *Eur. J. Nucl. Med.*, vol. 27, no. 9, pp. 1387–1398, 2000.
- [175] T. C. Chao and X. G. Xu, "Specific absorbed fractions from the image-based VIP-Man body model and EGS4-VLSI Monte Carlo code: Internal electron emitters," *Phys. Med. Biol.*, vol. 46, no. 4, pp. 901–927, 2001.
- [176] M. Zankl, N. Petoussi-Henss, U. Fill, and D. Regulla, "The application of voxel phantoms to the internal dosimetry of radionuclides," *Radiat. Protect. Dosimetry*, vol. 105, nos. 1–4, pp. 539–548, 2003.
- [177] M. Zankl, J. Becker, U. Fill, N. Petoussi-Henss, and K. F. Eckerman, "GSF male and female adult voxel models representing ICRP reference man—The present status," in *The Monte Carlo Method: Versatility Unbounded in a Dynamic Computing World*. Chattanooga, TN, USA: Versatility Unbounded Dyn. Comput. World, 2005.
- [178] International Commission on Radiological Protection, "Conversion coefficients for radiological protection quantities for external radiation exposures, Publication 116," *Ann. ICRP*, vol. 40, nos. 2–5, pp. 1–258, Apr. 2010.
- [179] International Commission on Radiological Protection, "The ICRP computational framework for internal dose assessment for reference adults: Specific absorbed fractions, Publication 133," *Ann. ICRP*, vol. 45, no. 2, pp. 5–73, 2016.
- [180] Y. S. Yeom, M. C. Han, C. H. Kim, and J. H. Jeong, "Conversion of ICRP male reference phantom to polygon-surface phantom," *Phys. Med. Biol.*, vol. 58, no. 19, pp. 6985–7007, 2013.
- [181] Y. Zhang, X. Li, W. P. Segars, and E. Samei, "Comparison of patient specific dose metrics between chest radiography, tomosynthesis, and CT for adult patients of wide ranging body habitus," *Med. Phys.*, vol. 41, no. 2, pp. 1–12, 2014.
- [182] FDA. *How FDA Plans to Help Consumers Capitalize on Advances in Science*. Accessed: Jul. 7, 2017. [Online]. Available: <https://www.fda.gov/NewsEvents/Newsroom/FDAVoices/ucm612016.htm>
- [183] T. M. Morrison, M. L. Dreher, S. Nagaraja, L. M. Angelone, and W. Kainz, "The role of computational modeling and simulation in the total product life cycle of peripheral vascular devices," *J. Med. Device*, vol. 11, no. 2, 2017, Art. no. 024503.
- [184] J. A. Nyenhuis *et al.*, "Heating near implanted medical devices by the MRI RF-magnetic field," *IEEE Trans. Mag.*, vol. 35, no. 5, pp. 4133–4135, Sep. 1999.
- [185] U.S. Food and Drug Administration (FDA). *Submission of Premarket Notifications for Magnetic Resonance Diagnostic Devices*. Accessed: Nov. 18, 2016. [Online]. Available: <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM454613.pdf>
- [186] U. S. Food and Drug Administration. *Assessment of Radiofrequency—Induced Heating in the Magnetic Resonance (MR) Environment for Multi-Configuration Passive Medical Devices*, Accessed: Mar. 22, 2016. [Online]. Available: <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM452644.pdf>
- [187] M. Murbach *et al.*, "Local SAR enhancements in anatomically correct children and adult models as a function of position within 1.5 T MR body coil," *Progr. Biophys. Mol. Biol.*, vol. 107, no. 3, pp. 428–433, 2011.
- [188] C. Li, W. P. Segars, G. D. Tourassi, J. M. Boone, and J. T. Dobbins, "Methodology for generating a 3D computerized breast phantom from empirical data," *Med. Phys.*, vol. 36, no. 7, pp. 3122–3131, 2009.
- [189] N. Kiarashi *et al.*, "Development of realistic physical breast phantoms matched to virtual breast phantoms based on human subject data," *Med. Phys.*, vol. 42, no. 7, pp. 4116–4126, 2015.
- [190] K. Bliznakova, Z. Bliznakov, V. Bravou, Z. Kolitsi, and N. Pallikarakis, "A three-dimensional breast software phantom for mammography simulation," *Phys. Med. Biol.*, vol. 48, no. 22, pp. 3699–3719, 2003.
- [191] P. R. Bakic, C. Zhang, and A. D. Maidment, "Development and characterization of an anthropomorphic breast software phantom based upon region-growing algorithm," *Med. Phys.*, vol. 38, no. 6, pp. 3165–3176, 2011.
- [192] A.-K. Carton, A. Grisey, P. M. de Carvalho, C. Dromain, and S. Muller, "A virtual human breast phantom using surface meshes and geometric internal structures," in *Proc. Int. Workshop Digit. Mammography*, 2014, pp. 356–363.
- [193] C. G. Graff, "A new open-source, multi-modality digital breast phantom," in *Proc. SPIE*, vol. 9783, 2016, Art. no. 978309.
- [194] G. M. Sturgeon, S. Park, W. P. Segars, and J. Y. Lo, "Synthetic breast phantoms from patient based eigenbreasts," *Med. Phys.*, vol. 44, no. 12, pp. 6270–6279, 2017.
- [195] X. Chen *et al.*, "High-resolution, anthropomorphic, computational breast phantom: Fusion of rule-based structures with patient-based anatomy," in *Proc. SPIE Med. Imag.*, Orlando, FL, USA, 2017, Art. no. 101321W.
- [196] E. Shaheen *et al.*, "The simulation of 3D mass models in 2D digital mammography and breast tomosynthesis," *Med. Phys.*, vol. 41, no. 8, pp. 1–17, 2014.
- [197] L. de Sisternes *et al.*, "A computational model to generate simulated three-dimensional breast masses," *Med. Phys.*, vol. 42, no. 2, pp. 1098–1118, 2015.
- [198] A. D. Maidment, "Virtual clinical trials for the assessment of novel breast screening modalities," in *Proc. Int. Workshop Digit. Mammography*, 2014, pp. 1–8.
- [199] A.-K. Carton *et al.*, "Development and validation of a simulation procedure to study the visibility of micro calcifications in digital mammograms," *Med. Phys.*, vol. 30, no. 8, pp. 2234–2240, 2003.
- [200] A. Maidment *et al.*, "OpenVCT: A framework for open-source virtual clinical trial resources," in *Proc. Amer. Assoc. Phys. Med. Ann. Meeting (AAPM)*, Denver, CO, USA, 2017, Art. no. 1057358.
- [201] A. Maidment *et al.* *The Open-Source Virtual Clinical Trial Project*. Accessed: Dec. 8, 2018. [Online]. Available: <https://sourceforge.net/projects/openvct/>
- [202] A. Badano *et al.*, "In silico imaging clinical trials for regulatory evaluation: Initial considerations for VICTRE, a demonstration study," in *Proc. SPIE Med. Imag.*, vol. 10132, 2017, pp. 1–6.
- [203] W. L. Oberkampf and C. J. Roy, *Verification and Validation in Scientific Computing*. New York, NY, USA: Cambridge Univ. Press, 2010.
- [204] P. Elangovan *et al.*, "Design and validation of realistic breast models for use in multiple alternative forced choice virtual clinical trials," *Phys. Med. Biol.*, vol. 62, no. 7, pp. 2778–2794, 2017.
- [205] A. Badano, "How much realism is needed?—The wrong question in silico imagers have been asking," *Med. Phys.*, vol. 44, no. 5, pp. 1607–1609, 2017.
- [206] *Qualification of Medical Device Development Tools (MDDT), Guidance for Industry, Tool Developers, and FDA Staff*. U. S. Food Drug Admin., Silver Spring, MD, USA, 2017.
- [207] R. K.-S. Kwan, A. C. Evans, and G. B. Pike, "MRI simulation-based evaluation of image-processing and classification methods," *IEEE Trans. Med. Imag.*, vol. 18, no. 11, pp. 1085–1097, Nov. 1999.
- [208] E. Neufeld, B. Lloyd, W. Kainz, and N. Kuster, "Functionalized anatomical models for computational life sciences," *Frontiers*, vol. 9, p. 1594, Oct. 2018.
- [209] Y. Zhang, Y. Feng, W. Wang, C. Yang, and P. Wang, "An expanded multi-scale Monte Carlo simulation method for personalized radiobiological effect estimation in radiotherapy: A feasibility study," *Sci. Rep.*, vol. 7, Mar. 2017, Art. no. 45019.
- [210] B. Zhang, J. Ma, L. Liu, and J. Cheng, "CNMAN: A Chinese adult male voxel phantom constructed from color photographs of a visible anatomical data set," *Radiat. Protect. Dosimetry*, vol. 124, no. 2, pp. 130–136, 2007.
- [211] K. Sato and F. Takahashi, "The contemporary JAEA Japanese voxel phantoms," *Radiat. Protect. Dosimetry*, vol. 149, no. 1, pp. 43–48, 2011.
- [212] GHBMC. (2017). *Elemance—Virtual Human Model Based Design and Protection Solutions*. [Online]. Available: <http://www.elemance.com/>